



Marea Professional Diving Eqpt

 CID9813 Opera V000050 Type:Vandagraph Trade

Marea Professional Diving Eqpt RN98, Les Quatre Chemins, La Garde, , 83130, France

Tel **0033 494762014** GBP Trade Price List Vandagraph (0009)



Credit 30 Days after date of invoice Credit Limit 0.00GBP

VAT FR79950013722

 Account Home

 Main Details

Account Memos

 Sales History

Purchase Orders

Stock References

 Cust./Supp. Returns

Service Visits

Distributor Agreements

Documents
Files

Linked Issues

Regulatory Agreements

Agreed to payment by CC upon dispatch of each order - RS 19/01/21

SRS68256

SRS Created : 09 Aug 2022

Created By Kate Griffiths

[Print Barcodes](#)

[Print A4 Sheet](#)

Split Items into Different SRS Numbers

Upload Document to SRS

Handle with care

Add Items to Supplier Returns Box

Add

☐ Return for Credit / Return to Stock
Unopened - will apply to all items on srs

☐ Clear From System - Force from System.
- will apply to all items on srs

Add notes to SRS

Add Notes

ADD DOCUMENT TO SRS : Document Description	Choose a file to upload: Choose file No file chosen
Next	

Service Repair Sheet 68256										
Company / Hospital		Marea Professional Diving Eqpt								
Contact Name		Francis Kalifa								
Department		Accounts								
Position		mareadiving@wanadoo.fr								
Company Phone		0033 494762014								
Direct Phone										
Opera Account		V000050								
Email		mareadiving@wanadoo.fr								
Order Number										
Viamed ORD Number										
Date Received Into building		22 Aug 2022								
Type		Quote Before Repair								
Issues?		272384								
Notes 09/Aug/2022 Kate Griffiths 22/Aug/2022 Robert Connor Received 1 x mixchek s/n 280373, 1 x VN202 s/n 02520. 06/Sep/2022 Catherine Spence engineers room LINKED ISSUE NOTES: 272384: AutoGenerated Issue SRS68256 Appears to be Complete										
Items for Repair										
... ..										
SRN	Equipment	Stock Reference	Barcode	Serialnumber	Description	Warranty New	Warranty Repair	Status	Repair Code	Time (H:M)

SRN35212	MixChek	MIXCHEK	 347979	280373	Mixchek	No	No	Complete - Repaired 20 Sep 2022 Michael Green	Display fault	01:30
SRN35212 Engineer Notes: INTERNAL NOTES not for customer discussion 02/Sep/2022 Michael Green The Mixchek analyser has been tested. The unit has a faulty Oxygen display. 06/Sep/2022 Michael Green The sensor has the date code H8. The gas sample tubing is discoloured. 20/Sep/2022 Michael Green The Mixchek analyser has been repaired. The Oxygen display and gas sample tubing have been replaced. The R-33S1 Oxygen sensor has been replaced with SN:83775. The supplied batteries have a 2018 expiry date and have not been reinstalled.										
SRN35213	Oxygen Analyser VN202	7910100	 1556038	02520	VN202mkII - Auto Switch Off	Yes	No	No Fault Found 20 Sep 2022 Michael Green	No Fault Found	00:00
SRN35213 Engineer Notes: INTERNAL NOTES not for customer discussion 02/Sep/2022 Michael Green The VN202 Oxygen analyser has been tested. No fault was found. A longer test will be performed. 20/Sep/2022 Michael Green The VN202 Oxygen analyser has been tested. No fault was found.										

Currently Viewing ITEM SRN35212 Oxygen Analyser VN202 7910100 02520 VN202mkII - Auto Switch Off

[Create a Meeting Issue](#)

Items Requested for Repair	Item notes	Engineer Notes	Notes To Customer
	The VN202 Oxygen analyser has been tested. No fault was found.	02/Sep/2022 Michael Green The VN202 Oxygen analyser has been tested. No fault was found. A longer test will be performed. 20/Sep/2022 Michael Green The VN202 Oxygen analyser has been tested. No fault was found.	The VN202 Oxygen analyser has been tested. No fault was found.