



Kate Griffiths <viamed.kate.griffiths@gmail.com>

Re: Order 27042273ZT1 Tracking 2898611962

1 message

Sudath Weeraman <sudathweeraman17@gmail.com>

21 September 2022 at 16:25

To: kate.griffiths@viamed.co.uk

Dear Kate,

Thanks for your email. Sorry for the late reply,

We suggest order 1 pc of R-49V Oxygen sensor for test PB 840.

Can you send proforma invoice for above sensor..

Thanks & Regards

Sudath Weeraman
Prime Lanka Associates
Sri Lanka.

On Fri, Sep 16, 2022, 7:03 PM Kate Griffiths <office@viamed.co.uk> wrote:

Hi Sudath,

Ryan Swaine has advised me that if you are using the R-24MED, then we can replace it with the R-49V. We can guarantee that the R-49V is a direct replacement for the R-24MED, but we can not guarantee it is for the PB840.

I am sorry I am unable to be of more assistance.

Kind regards

Kate Griffiths

Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.

<http://www.viamed.co.uk>Email kate.griffiths@viamed.co.uk

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd

Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

On Thu, 15 Sept 2022 at 16:34, Sudath Weeraman <sudathweeraman17@gmail.com> wrote:

Dear Kate,

The BP 840 ventilator currently use
R-24MED Oxygen sensor.

We are supplied many times

Best regards,

Sudarth Weeraman
Prime Lanka Associates
Sri Lanka.

On Thu, Sep 15, 2022, 8:49 PM Kate Griffiths <office@viamed.co.uk> wrote:

Hi Sudath,

There are a variety of sensors that may be compatible with the PB 840. Are you able to let us know which sensor is currently being used in the PB 840 and then we can make sure we offer the correct sensor?

Many thanks.

Kind regards

Kate Griffiths

Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.

<http://www.viamed.co.uk>
Email kate.griffiths@viamed.co.uk
Tel: +44 (0)1535 634542
Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd
Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy
The information in this email is confidential and may be legally privileged.
It is intended solely for the addressee. Access to this email by anyone else is unauthorized.
If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

On Thu, 15 Sept 2022 at 15:45, Sudath Weeraman <sudathweeraman17@gmail.com> wrote:

Dear Kate,

Okay..not any compatible sensor for PB 840?

Best regards,

Sudath Weeraman
Prime Lanka Associates
Sri Lanka.

On Thu, Sep 15, 2022, 7:57 PM Kate Griffiths <office@viamed.co.uk> wrote:

Good afternoon Sudath,

The R-49V oxygen sensor is **not** compatible for Puritan Bennett 840 Ventilator (PB-840).

Many thanks.

Kind regards

Kate Griffiths

Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.

<http://www.viamed.co.uk>

Email kate.griffiths@viamed.co.uk
Tel: +44 (0)1535 634542
Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd
Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

On Thu, 15 Sept 2022 at 14:49, Sudath Weeraman <sudathweeraman17@gmail.com> wrote:

Dear Kate,

Thank you for your email,

R-49V oxygen sensor compatible for Puratant bennete 840 Ventilator?
(PB-840)

Best regards

Sudath Weeraman
Prime Lanka Associates
Sri Lanka.

On Thu, Sep 15, 2022, 3:45 PM Kate Griffiths <office@viamed.co.uk> wrote:

Good morning Sudath,

The R-49V oxygen sensor is manufactured for Viamed in Germany by Envitech.

Many thanks.

Kind regards

Kate Griffiths

Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.

<http://www.viamed.co.uk>

Email kate.griffiths@viamed.co.uk

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd

Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

On Wed, 14 Sept 2022 at 16:39, Sudath Weeraman <sudathweeraman17@gmail.com> wrote:

Dear Kate,

Thank you for your reply, R-49V Oxygen sensor manufacturer in Viamed?

We will send new quotation for 15 Pcs to my client and get the order

Best regards,

Sudath Weeraman
Prime Lanka Associates
Sri Lanka.

On Wed, Sep 14, 2022, 8:28 PM Kate Griffiths <office@viamed.co.uk> wrote:

Good afternoon Sudath,

Thank you for your payment.

The R-24MED oxygen sensor, part number 0110024 is still not in stock, we are still experiencing supply issues and do not have an accurate lead time.

An alternative is R-49V, part number 0110049. This item is currently in stock and costs £34 each EXW.

Many thanks.

Kind regards

Kate Griffiths

Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.

<http://www.viamed.co.uk>

Email kate.griffiths@viamed.co.uk

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd

Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

On Wed, 14 Sept 2022 at 14:04, Sudath Weeraman <sudathweeraman17@gmail.com> wrote:

Dear Kate,

Good Morning

Settled for RVM138476-1 / GBP-597.84, already informed to account director Ms.Helen
Need to pay RVM138974-1 / GBP-1108.76 ,will arrange immediately,

How is the R-24MED Oxygen sensor? Is it available next week?