

Viamed Returns Shipping Invoice

Delivery Address:

Honeywell Healthcare Solutions GmbH RMA: 300279792

Andre Schliemann

Alter Holzhafen 18

Wismar

Germany

D-23966

Eori DE3174778

0049 3841 360204

Date: 11 Oct 2024

Viamed RN: RTS11001000

Shipper Address:

Viamed

15 Station Road

Cross Hills

Keighley

West Yorkshire

BD20 7DT

+44 (0) 1535 634542

+44 (0) 1535 635582

info@viamed.co.uk

Returns Under Warranty Value for Customs : 5.00 Euro

See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
0110043	0110043	E1002780 OOM112 Oxygen Sensor	1
	Tariff Code: 9019200000		

Part Number	Serial Number	RMA	Fault Description
0110043	V110524/V110481	300279792	Manufacturing defect, loose PCB. Slip ring contacts are not properly attached to cell housing. Matched pair Country of Origin: Germany Custom Value: 5.00 euro