

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Cynar Magno
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 27 May 2025
RMA:
Viamed RN: RTS11001022

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Part Number	P/N / Value	Description	Quantity
C44611-R22AHJR	0110361 (\$27.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Oxygen sensor R-22AHJR Jikco	22
C44611-R22A	0110122 (\$20.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	1

Part Number	Serial Number	RMA	Fault Description
C44611-R22A	222857		Output (7mV - 13mV) 2.7 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	134826		Output (10.5 - 13.5 mV) 9.9mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134828		Output (10.5 - 13.5 mV) 10.2 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134830		Output (10.5 - 13.5 mV) 10.2 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A156038		Failed 10.3mV failed in QA did not reach end user
C44611-R22AHJR	A221697		Output (10.5 - 13.5 mV) 9.6 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A221705		Output (10.5 - 13.5 mV) 10.4 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A221707		Output (10.5 - 13.5 mV) 9.4 m V Failed in Q.A. Did not reach end user
C44611-R22AHJR	A221713		Output (10.5 - 13.5 mV) 10.3mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A221718		Output (10.5 - 13.5 mV) 10.1 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239149		Output (10.5 - 13.5 mV) 14.0 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239154		Output (10.5 - 13.5 mV) 10.4 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239173		Output (10.5 - 13.5 mV) 10.4 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239180		Output (10.5 - 13.5 mV) 10.0 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239184		Output (10.5 - 13.5 mV) 10.1 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239196		Output (10.5 - 13.5 mV) 10.4 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239199		Output (10.5 - 13.5 mV) 10.1 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239207		Output (10.5 - 13.5 mV) 10.3 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239231		Output (10.5 - 13.5 mV) 10.1mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239235		Output (10.5 - 13.5 mV) 10.3 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239241		Output (10.5 - 13.5 mV) 10.0 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239243		Output (10.5 - 13.5 mV) 10.3 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A239244		Output (10.5 - 13.5 mV) 10.0 mV Failed in Q.A. Did not reach end user