

Viamed Returns Shipping Invoice

Delivery Address:
Maxtec Inc
Keely Mitchell
2305 South 1070 West
Salt Lake City
Utah, USA
84119
0

Date: 23 Jul 2025
RMA: 353628
Viamed RN: RTS11001037

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$934.76
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
R218P12-001	0111250 Tariff Code: 90181990-00	Maxtec Oxygen Analyser with in	1
R217P72	0111261 Tariff Code: 90181990-00 Country of Origin: USA	Maxtec Oxygen Analyser - MaxO2	1
R230P01-001	0111263 Tariff Code: 90181990-00	Maxtec Oxygen Monitor - MaxO2	1

Part Number	Serial Number	RMA	Fault Description
R218P12-001	LC64999056	353628	Failed in Q.A. Did not reach end user Calibrate button not working. Missing Parts No lanyard, adapter, flow diverter and tee adapter Country of Origin: USA Customs value: \$151.90
R217P72	LB25399047	353628	Failed in Q.A. Did not reach end user On/Off button is intermittent Country of Origin: USA Customs value: \$303.79
R230P01-001	JJ60899011	353628	Display fault Failed in service Country of Origin: USA Customs value: \$479.07