

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments

16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 30 Aug 2025
RMA:
Viamed RN: RTS11001047

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Part Number	P/N / Value	Description	Quantity
C44611-R22AV	0110132 (\$18.00)	Automotive Oxygen Sensor R-22A	3
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
C44611-R22A	0110122 (\$20.00)	Automotive Oxygen Sensor R-22A	3
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		

Part Number	Serial Number	RMA	Fault Description
C44611-R22A	278511		Output (7mV - 13mV) 2.1 n Failed in Q.A. Did not reach end user
C44611-R22A	278514		Output (7mV - 13mV) 0 mV Failed in Q.A. Did not reach end user
C44611-R22A	278517		Output (7mV - 13mV) 3.3 mV Failed in Q.A. Did not reach end user
C44611-R22AV	269587		Output (9mV - 13mV) 15.8 mV Failed in Q.A. Did not reach end user
C44611-R22AV	277489		Output (9mV - 13mV) 00.1 mV Failed in Q.A. Did not reach end user
C44611-R22AV	277901		Output (9mV - 13mV) 00.1 mV Failed in Q.A. Did not reach end user