

# Viamed Returns Shipping Invoice

Delivery Address:  
Teledyne Analytical Instruments  
  
16830 Chestnut Street  
City Of Industry  
California  
91748-1020  
0016269341589

Date: 02 Oct 2025  
RMA:  
Viamed RN: RTS11001049

Shipper Address:  
Viamed  
15 Station Road  
Cross Hills  
Keighley  
West Yorkshire  
BD20 7DT  
+44 (0) 1535 634542  
+44 (0) 1535 635582  
info@viamed.co.uk

---

| Part Number   | P/N / Value                                                                 | Description                    | Quantity |
|---------------|-----------------------------------------------------------------------------|--------------------------------|----------|
| C-73894       | 0110081 (\$110.00)<br>Tariff Code: 9027.10.2000                             | Teledyne Oxygen sensor UFO 130 | 1        |
| C43690-R17MED | 0110017 (\$27.00 Ea)<br>Tariff Code: 9027.10.2000<br>Country of Origin: USA | Teledyne Oxygen sensor R-17MED | 1        |

| Part Number   | Serial Number | RMA | Fault Description  |
|---------------|---------------|-----|--------------------|
| C43690-R17MED | 259815        |     | Unstable           |
| C-73894       | 202153        |     | Calibration Failed |

| Part Number   | Serial No. | RMA | Fault Report |
|---------------|------------|-----|--------------|
| C43690-R17MED | 259815     |     |              |
| C-73894       | 202153     |     |              |