

Viamed Returns Shipping Invoice

Delivery Address:

Bluepoint Medical GmbH & Co. KG (Europe)

Anja Merz

An der Trave 15

Selmsdorf

Germany

23923

0049 38823 5488020

Date: 18 Jan 2026

EMA:

Viamed RN: RTS11001064

Shipper Address:

Viamed

15 Station Road

Cross Hills

Keighley

West Yorkshire

BD20 7DT

+44 (0) 1535 634542

+44 (0) 1535 635582

info@viamed.co.uk

Part Number

3090112007

P/N / Value

4410520

Tariff Code: 9018199000

Description

Sidestream Capnograph - VM-250

Quantity

1

Part Number	Serial Number	RMA	Fault Description
3090112007	A1711160013		Calibration test, Co2 percentage is off even after zeroing. Quote needed before proceeding

Part Number
3090112007

Serial No. RMA
A1711160013

Fault Report