

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Warranty Returns
16830 Chestnut Street
City Of Industry
California
91748-1020

Date: 26 Feb 2026
RMA: 0062868
Viamed RN: RTS11001071

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$99
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$33.00 Ea) Tariff Code: 9027.10.2000 Country of Origin: USA	Teledyne Oxygen sensor R-17MED	3

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	326404	0062868	Output (7mV - 13mV) 00.1mV Failed in Q.A. Did not reach end user Country of Origin: USA Custom Value: \$33.00
C43690-R17MED	326440	0062868	Output (7mV - 13mV) 8.8-00.1 mV Failed in Q.A. Did not reach end user Country of Origin: USA Custom Value: \$33.00
C43690-R17MED	326453	0062868	Output (7mV - 13mV) 00.1-10.8 erratic mV Failed in Q.A. Did not reach end user Country of Origin: USA Custom Value: \$33.00

Part Number	Serial No.	RMA	Fault Report
C43690-R17MED	326404	0062868	
C43690-R17MED	326440	0062868	
C43690-R17MED	326453	0062868	