

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Repairs
16830 Chestnut Street
City Of Industry
California
91748-1017

Date: 31 Mar 2026
RMA:
Viamed RN: RTS11001076

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
0110127	0110127 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Oxygen sensor - Automotive R-	15

Part Number	Serial Number	RMA	Fault Description
0110127	340382		Output (9mV - 13mV) 13.7mV Failed in Q.A. Did not reach end user
0110127	340388		Output (9mV - 13mV) 14.0mV Failed in Q.A. Did not reach end user
0110127	340389		Output (9mV - 13mV) 13.3 mV Failed in Q.A. Did not reach end user
0110127	340390		Output (9mV - 13mV) 13.2mV Failed in Q.A. Did not reach end user
0110127	340391		Output (9mV - 13mV) 13.4 mV Failed in Q.A. Did not reach end user
0110127	340393		Output (9mV - 13mV) 15.0 mV Failed in Q.A. Did not reach end user
0110127	340401		Output (9mV - 13mV) 14.2 mV Failed in Q.A. Did not reach end user
0110127	340402		Output (9mV - 13mV) 13.9 mV Failed in Q.A. Did not reach end user
0110127	340405		Output (9mV - 13mV) 14.6 mV Failed in Q.A. Did not reach end user
0110127	340416		Output (9mV - 13mV) 13.3 mV Failed in Q.A. Did not reach end user
0110127	340417		Output (9mV - 13mV) 13.5mV Failed in Q.A. Did not reach end user
0110127	340418		Output (9mV - 13mV) 13.4 mV Failed in Q.A. Did not reach end user
0110127	340425		Output (9mV - 13mV) 14.3 mV Failed in Q.A. Did not reach end user
0110127	340425		Output (9mV - 13mV) 14.3 mV Failed in Q.A. Did not reach end user
0110127	340428		Output (9mV - 13mV) 13.5 mV Failed in Q.A. Did not reach end user

Part Number	Serial No.	RMA	Fault Report
0110127	340382		
0110127	340388		
0110127	340389		
0110127	340390		
0110127	340391		
0110127	340393		
0110127	340401		
0110127	340402		
0110127	340405		
0110127	340416		
0110127	340417		
0110127	340418		
0110127	340425		
0110127	340425		
0110127	340428		