

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Warranty Returns
16830 Chestnut Street
City Of Industry
California
91748-1017
0016269341589

Date: 16 May 2026
RMA:
Viamed RN: RTS11001083

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Part Number	P/N / Value	Description	Quantity
C43690-R17AH	0110137 (\$33.00 Ea) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	2
C44611-R22AV	0110132 (\$20.00 ea) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	2
C44611-R22A	0110122 (\$17.00 ea) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	2

Part Number	Serial Number	RMA	Fault Description
C44611-R22A	333802		Output (7mV - 13mV) 2.7 mV Failed in Q.A. Did not reach end user
C44611-R22A	333899		Output (7mV - 13mV) 00.0 mV Failed in Q.A. Did not reach end user
C44611-R22AV	328601		Output (9mV - 13mV) 8.8 mV Failed in Q.A. Did not reach end user
C44611-R22AV	342249		Output (9mV - 13mV) 00.2 mV Failed in Q.A. Did not reach end user
C43690-R17AH	355234		Output (10.5mV - 13.5mV) 13.6 mV Failed in Q.A. Did not reach end user
C43690-R17AH	355302		Output (10.5mV - 13.5mV) 14.4 mV Failed in Q.A. Did not reach end user

Part Number	Serial No.	RMA	Fault Report
C44611-R22A	333802		
C44611-R22A	333899		
C44611-R22AV	328601		
C44611-R22AV	342249		
C43690-R17AH	355234		
C43690-R17AH	355302		