

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Warranty Returns
16830 Chestnut Street
City Of Industry
California
91748-1017
0016269341589

Date: 23 Jul 2026
RMA: 0063547
Viamed RN: RTS11001090

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$249.00

See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
A37016-T1	0110051 (\$33.00 Ea) Tariff Code: 9027.10.2000 Country of Origin: USA	Teledyne Oxygen sensor T-1	4
C-73894	0110081 (\$117.00 each) Tariff Code: 9027.10.2000	Teledyne Oxygen sensor UFO 130	1

Part Number	Serial Number	RMA	Fault Description
A37016-T1	338794	0063547	Leaking Failed in Q.A. Did not reach end user
A37016-T1	338795	0063547	Leaking Failed in Q.A. Did not reach end user
A37016-T1	338798	0063547	Leaking Failed in Q.A. Did not reach end user
A37016-T1	338799	0063547	Leaking Failed in Q.A. Did not reach end user
C-73894	348706	0063547	leaking Failed in Q.A. Did not reach end user

Part Number	Serial No.	RMA	Fault Report
A37016-T1	338794	0063547	
A37016-T1	338795	0063547	
A37016-T1	338798	0063547	
A37016-T1	338799	0063547	
C-73894	348706	0063547	