

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Warranty Returns
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 10 Aug 2026
RMA:
Viamed RN: RTS11001092

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Part Number	P/N / Value	Description	Quantity
C44611-R22AHJR	0110361 (\$20.00 ea) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Oxygen sensor R-22AHJR Jikco	9
C43690-R17AH	0110137 (\$33.00 Ea) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	1

Part Number	Serial Number	RMA	Fault Description
C43690-R17AH	376185		Output (10.5mV - 13.5mV) 00.1 Failed in Q.A. Did not reach end user
C44611-R22AHJR	A377307		Output (10.5 - 13.5 mV) 10.0 Failed in Q.A. Did not reach end user
C44611-R22AHJR	A377324		Output (10.5 - 13.5 mV) 9.5 Failed in Q.A. Did not reach end user
C44611-R22AHJR	A377331		Output (10.5 - 13.5 mV) 9.9 Failed in Q.A. Did not reach end user
C44611-R22AHJR	A377334		Output (10.5 - 13.5 mV) 10.0 Failed in Q.A. Did not reach end user
C44611-R22AHJR	A377345		Output (10.5 - 13.5 mV) 10.0 Failed in Q.A. Did not reach end user
C44611-R22AHJR	A377365		Output (10.5 - 13.5 mV) 9.7 Failed in Q.A. Did not reach end user
C44611-R22AHJR	A377377		Output (10.5 - 13.5 mV) 9.6 Failed in Q.A. Did not reach end user
C44611-R22AHJR	A377446		Output (10.5 - 13.5 mV) 10.0 Failed in Q.A. Did not reach end user
C44611-R22AHJR	A377456		Output (10.5 - 13.5 mV) 9.6 Failed in Q.A. Did not reach end user

Part Number	Serial No.	RMA	Fault Report
C43690-R17AH	376185		
C44611-R22AHJR	A377307		
C44611-R22AHJR	A377324		
C44611-R22AHJR	A377331		
C44611-R22AHJR	A377334		
C44611-R22AHJR	A377345		
C44611-R22AHJR	A377365		
C44611-R22AHJR	A377377		
C44611-R22AHJR	A377446		
C44611-R22AHJR	A377456		