

Viamed Returns Shipping Invoice

Delivery Address:
ChoiceMMed Germany GmbH
Bill Nie
Wiesenstrasse 21
Dusseldorf
Germany
40549
0049 211 5065 9866

Date: 16 Sep 2025
RMA:
Viamed RN: RTS1100811

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00

See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
MD300C2	2810011 Tariff Code: 90181990-00	MD300-C2 (MD300-D) OLED Finger	8
MD300C15D G/W	2810046 Tariff Code: 90181990-00	MD300-C15D LCD Finger Pulse Ox	1
MD300C29	2810050 Tariff Code: 90181990-00	MD300-C29 Finger Pulse Oximete	1
2810055	2810055 Tariff Code: 90181990-00	A5 OLED Finger Pulse Oximeter	4
2810052	2810052	A2 LED Finger Pulse Oximeter	1

Part Number	Serial Number	RMA	Fault Description
MD300C2	190909001821		keeps switching off
MD300C2	190909001827		keeps switching off
MD300C2	190909002877		Keeps switching off
MD300C2	190909004926		Keeps switching off
MD300C2	190909004943		Keeps switching off
MD300C2	190909004952		Keeps switching off
MD300C2	190909004955		Keeps switching off
MD300C2	190909007608		Will not stay on
MD300C15D G/W	193208004903		Numbers Erratic, numbers keep fading
MD300C29	202569976543		No Output
2810052	NFDD421787		Hinges Broken
2810055	CCAR172917		Display in not central
2810055	CCAR172931		Display not central
2810055	CCAR172933		Display is not central
2810055	CCAR172967		Display is not central