

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 28 Feb 2023
RMA:
Viamed RN: RTS1100884

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17A-LV	0110107 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	29
C43690-R17AH	0110137 (\$33.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	3
C43690-R17AV	0110127 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	1
C44611-R22AHJR	0110361 (\$27.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Oxygen sensor R-22AHJR Jikco	22
C44611-R22A	0110122 (\$20.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	2

Part Number	Serial Number	RMA	Fault Description
C43690-R17A-LV	843983		Output (7mv - 11.5mv) 11.8 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843984		Output (7mv - 11.5mv) 12.5 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843988		Output (7mv - 11.5mv) 12.0 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843989		Output (7mv - 11.5mv) 2.2 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843994		Output (7mv - 11.5mv) 12.2 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843996		Output (7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843997		Output (7mv - 11.5mv) 12.0 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843999		Output (7mv - 11.5mv) 12.2 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844000		Output (7mv - 11.5mv) 11.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844002		Output (7mv - 11.5mv) 12.0 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844005		Output (7mv - 11.5mv) 11.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844008		Output (7mv - 11.5mv) 11.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844011		Output (7mv - 11.5mv) 00.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844012		Pass Yes n Move Item to location 126541 n Output (7mv - 11.5mv) 12.0 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844017		Output (7mv - 11.5mv) 12.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844019		Output (7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844022		Output (7mv - 11.5mv) 2.8 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844029		output 12.4mV failed qa did not reach end user
C43690-R17A-LV	844037		Output (7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844038		Output (7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844039		Output (7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844045		Output (7mv - 11.5mv) 12.2 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844047		Output (7mv - 11.5mv) 12.0 n Move Item to location 126541 n Output (7mv - 11.5mv) 11.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844049		Output (7mv - 11.5mv) 11.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844055		Output (7mv - 11.5mv) 11.6 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844060		Output (7mv - 11.5mv) 12.3 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844065		Output (7mv - 11.5mv) 12.3 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844072		Output (7mv - 11.5mv) 11.8 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user

Part Number	Serial Number	RMA	Fault Description
C43690-R17A-LV	844073		Output (7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C44611-R22A	824470		Output (7mV - 13mV) 4.5 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
C44611-R22A	832094		Output (7mV - 13mV) 00.1 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
C43690-R17AV	844120		Output (9mV - 13mV) 00.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17AH	843753		Output (10.5mV - 13.5mV) 00.6 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17AH	843781		Output (10.5mV - 13.5mV) 1.6 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17AH	843818		Output (10.5mV - 13.5mV) 1.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838455		Output (10.5 - 13.5 mV) 14.3 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838459		Output (10.5 - 13.5 mV) 14.0 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838464		Output (10.5 - 13.5 mV) 14.5 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838465		Output (10.5 - 13.5 mV) 15.4 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838466		Output (10.5 - 13.5 mV) 13.6 pressure 997 MB n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838491		Output (10.5 - 13.5 mV) 13.5 n Pass Yes n 126550 n Output (10.5 - 13.5 mV) 13.7 Pressure 997 Mb n Pass Yes n 126550 n Output (10.5 - 13.5 mV) 13.7 Pressure 997 Mb n 126550 n Output (10.5 - 13.5 mV) 13.7 Pressure 997 Mb n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838500		Output (10.5 - 13.5 mV) 15.3 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838505		Output (10.5 - 13.5 mV) 14.0 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838506		Output (10.5 - 13.5 mV) 13.8 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838517		Output (10.5 - 13.5 mV) 14.0 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838521		Output (10.5 - 13.5 mV) 13.8 n Pass Yes n 126550 n Output (10.5 - 13.5 mV) 13.8 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838535		Output (10.5 - 13.5 mV) 13.9 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838537		Output (10.5 - 13.5 mV) 13.7 Pressure 997 MB n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838540		Output (10.5 - 13.5 mV) 15.0 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838547		Output (10.5 - 13.5 mV) 14.4 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838564		Output (10.5 - 13.5 mV) 14.5 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838575		Output (10.5 - 13.5 mV) 15.5 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838595		Output (10.5 - 13.5 mV) 15.0 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838620		Output (10.5 - 13.5 mV) 13.7 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838632		Output (10.5 - 13.5 mV) 13.5 Pressure 997 Mb n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838636		Output (10.5 - 13.5 mV) 13.4 Pressure 997 Mb n 126550 n Failed in Q.A. n Did not reach end user

Part Number	Serial Number	RMA	Fault Description
C44611-R22AHJR	A838638		Output (10.5 - 13.5 mV) 13.8 n 126550 n Failed in Q.A. n Did not reach end user