

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 28 Aug 2025
RMA:
Viamed RN: RTS1100891

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$671.69
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C44611-R22MED	0110022 (\$25.00 Ea.) Tariff Code: 9027.10.2000 Country of Origin: USA	Teledyne Oxygen sensor R-22MED	3
C-73894	0110081 (\$110.00) Tariff Code: 9027.10.2000	Teledyne Oxygen sensor UFO 130	5
A51327-T7	0110057 (\$46.69 Ea.) Tariff Code: 9019200000 Country of Origin: USA	Teledyne Oxygen sensor T-7	1

Part Number	Serial Number	RMA	Fault Description
C44611-R22MED	618876		Non Linear Country of Origin: USA Customs value. \$25.00
C44611-R22MED	624838		unstable and nonlinear Country of Origin: USA Customs value. \$25.00
C44611-R22MED	624844		unstable and nonlinear Country of Origin: USA Customs value. \$25.00
A51327-T7	636166		Failed Country of Origin: USA Customs value. \$46.69
C-73894	479035		Failed
C-73894	500812		Failed
C-73894	689239		Unstable output signal Country of Origin: USA Customs value. \$110.00
C-73894	689248		Unstable output signal Country of Origin: USA Customs value. \$110.00
C-73894	689251		Unstable output signal Country of Origin: USA Customs value. \$110.00