

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street

Date: 09 Apr 2025
RMA:
Viamed RN: RTS1100902

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

91748-1020

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C44611-R22A	0110122 (\$20.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	7
C44611-R22AV	0110132 (\$18.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	3
C43690-R17A	0110117 (\$27.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	8
C43690-R17A-LV	0110107 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	2
C43981-R21A	0110121 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-21A	1
C74069-R33S	0110214 Tariff Code: HTS Code 9027.90.58.10	R-33S Oxygen Sensor	1
C74069-R33S1	0110215 Tariff Code: HTS Code 9027.90.58.10	R-33S1 Oxygen Sensor	1

Part Number	Serial Number	RMA	Fault Description
C43690-R17A-LV	865805		Output (7mv - 11.5mv) 00.1 mV Erratic Failed in Q.A. Did not reach end user
C43690-R17A-LV	865841		Output (7mv - 11.5mv) 13.8mV Failed in Q.A. Did not reach end user
C43690-R17A	778219		Output (7mV - 13mV) 00.2mV Failed in Q.A. Did not reach end user
C43690-R17A	778243		Output (7mV - 13mV) erratic mV Failed in Q.A. Did not reach end user
C43690-R17A	778267		Output (7mV - 13mV) 13.8 mV Failed in Q.A. Did not reach end user
C43690-R17A	778269		Output (7mV - 13mV) 9.9mV Failed in Q.A. Did not reach end user
C43690-R17A	778278		Output (7mV - 13mV) 01.5mV Failed in Q.A. Did not reach end user
C43690-R17A	778301		Output (7mV - 13mV) 00.1mV Failed in Q.A. Did not reach end user
C43690-R17A	812619		Output (7mV - 13mV) 14.1mV Failed in Q.A. Did not reach end user
C43690-R17A	812622		Output (7mV - 13mV) 14.0mV Failed in Q.A. Did not reach end user
C43981-R21A	898918		Output (9mV - 14mV) 15.7mV Failed in Q.A. Did not reach end user
C44611-R22A	867257		Output (7mV - 13mV) 00.1mV Failed in Q.A. Did not reach end user
C44611-R22A	867401		Output (7mV - 13mV) 00.1mV Failed in Q.A. Did not reach end user
C44611-R22A	867429		Output (7mV - 13mV) 00.1mV Failed in Q.A. Did not reach end user
C44611-R22A	871471		Output (7mV - 13mV) 47.1mV Failed in Q.A. Did not reach end user
C44611-R22A	871590		Output (7mV - 13mV) 1.5 mV Failed in Q.A. Did not reach end user
C44611-R22A	880798		Output (7mV - 13mV) 16.3 mV Failed in Q.A. Did not reach end user
C44611-R22A	880852		Output (7mV - 13mV) 00.3 mV Failed in Q.A. Did not reach end user
C44611-R22AV	868433		Output (9mV - 13mV) 14.1 mV Failed in Q.A. Did not reach end user
C44611-R22AV	914324		Output (9mV - 13mV) 13.3mVFailed in Q.A. Did not reach end user
C44611-R22AV	914365		Output (9mV - 13mV) 14.8mV Failed in Q.A. Did not reach end user
C74069-R33S	829619		failed QA unstable output when tipped Failed in Q.A. Did not reach end user
C74069-R33S1	869051		High output Failed in Q.A. Did not reach end user