

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 08 Nov 2024
RMA:
Viamed RN: RTS1100913

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17A-LV	0110107 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	1
C43690-R17A	0110117 (\$27.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	1
C43690-R17AH	0110137 (\$33.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	6
C43690-R17AV	0110127 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	2
C44611-R22A	0110122 (\$20.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	1
C43981-R21A	0110121 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-21A	1

Part Number	Serial Number	RMA	Fault Description
C43690-R17A-LV	933517		Output (7mv - 11.5mv) 12.7mV in Q.A. Did not reach end user
C43690-R17A	778362		Output (7mV - 13mV) 13.2mV Failed in Q.A. Did not reach end user
C43981-R21A	934222		Output 9mV - 14mV 8.4 mV Failed in Q.A. Did not reach end user
C44611-R22A	931572		Output (7mV - 13mV) 6.1 mV Failed in Q.A. Did not reach end user
C43690-R17AV	933558		Output (9mV - 13mV) 13.4 mV Failed in Q.A. Did not reach end user
C43690-R17AV	933589		Output (9mV - 13mV) 13.4 mV Failed in Q.A. Did not reach end user
C43690-R17AH	933412		Output (10.5mV - 13.5mV) 15.0 mV Failed in Q.A. Did not reach end user
C43690-R17AH	933459		Output (10.5mV - 13.5mV) 14.1 mV Failed in Q.A. Did not reach end user
C43690-R17AH	933464		Output (10.5mV - 13.5mV) 13.9 mV Failed in Q.A. Did not reach end user
C43690-R17AH	933473		Output (10.5mV - 13.5mV) 13.9 mV Failed in Q.A. Did not reach end user
C43690-R17AH	933478		Output (10.5mV - 13.5mV) 14.3 mV Failed in Q.A. Did not reach end user
C43690-R17AH	933488		Output (10.5mV - 13.5mV) 13.7 mV Failed in Q.A. Did not reach end user