

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 10 Feb 2025
RMA:
Viamed RN: RTS1100914

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$27.00 Ea) Tariff Code: 9027.10.2000 Country of Origin: USA	Teledyne Oxygen sensor R-17MED	3
C44611-R22MED	0110022 (\$25.00 Ea.) Tariff Code: 9027.10.2000 Country of Origin: USA	Teledyne Oxygen sensor R-22MED	1
C-73894	0110081 (\$110.00) Tariff Code: 9027.10.2000	Teledyne Oxygen sensor UFO 130	1

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	917068		Output (7mV - 13mV) 00.1 mV Failed in Q.A. Did not reach end user
C43690-R17MED	917092		Output (7mV - 13mV) 00.1 mV Failed in Q.A. Did not reach end user
C43690-R17MED	917419		Output (7mV - 13mV) erratic 00.1-9.6 mV Failed in Q.A. Did not reach end user
C44611-R22MED	909271		Output (7mV - 13mV) 00.1 mV Failed in Q.A. Did not reach end user
C-73894	905969		Failed in QA. Did not reach end user