

# Viamed Returns Shipping Invoice

Delivery Address:  
Teledyne Analytical Instruments  
Margie Read  
16830 Chestnut Street  
City Of Industry  
California  
91748-1020  
0016269341589

Date: 27 May 2025  
RMA:  
Viamed RN: RTS1100933

Shipper Address:  
Viamed  
15 Station Road  
Cross Hills  
Keighley  
West Yorkshire  
BD20 7DT  
+44 (0) 1535 634542  
+44 (0) 1535 635582  
info@viamed.co.uk

---

Part Number	P/N / Value	Description	Quantity
C44611-R22MED	0110022 (\$25.00 Ea.) Tariff Code: 9027.10.2000 Country of Origin: USA	Teledyne Oxygen sensor R-22MED	6

Part Number	Serial Number	RMA	Fault Description
C44611-R22MED	956283		Output ( 7mV - 13mV ) 15.1 Failed in Q.A. Did not reach end user
C44611-R22MED	956286		Output ( 7mV - 13mV ) 15.9 Failed in Q.A. Did not reach end user
C44611-R22MED	956313		Output ( 7mV - 13mV ) 14.9 Failed in Q.A. Did not reach end user
C44611-R22MED	956333		Output ( 7mV - 13mV ) 14.2 Failed in Q.A. Did not reach end user
C44611-R22MED	956461		Output ( 7mV - 13mV ) 15.7 Failed in Q.A. Did not reach end user
C44611-R22MED	956474		Output ( 7mV - 13mV ) 16.4 Failed in Q.A. Did not reach end user