

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 27 May 2025
RMA:
Viamed RN: RTS1100943

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Part Number	P/N / Value	Description	Quantity
C44611-R22A	0110122 (\$20.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	6
C44611-R22AV	0110132 (\$18.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	3
C44611-R22AHJR	0110361 (\$27.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Oxygen sensor R-22AHJR Jikco	67
C43690-R17AH	0110137 (\$33.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	5

Part Number	Serial Number	RMA	Fault Description
C44611-R22A	985089		Output (7mV - 13mV) 21.6mV Failed in Q.A. Did not reach end user
C44611-R22A	986592		Output (7mV - 13mV) 2.4 mV Failed in Q.A. Did not reach end user
C44611-R22A	994290		Output (7mV - 13mV) 0.2-6.4 mV Failed in Q.A. Did not reach end user
C44611-R22A	994290		Output (7mV - 13mV) 0.2-6.4 mV Failed in Q.A. Did not reach end user
C44611-R22A	994303		Output (7mV - 13mV) 00.1 n Move Item to location 126552 n Test Equipment Barcode 126552 n Failed in Q.A. Did not reach end user
C44611-R22A	994429		Output (7mV - 13mV) 48.8 n Move Item to location 126552 n Test Equipment Barcode 126552 n Failed in Q.A. Did not reach end user
C44611-R22AV	100029		Output (9mV - 13mV) one of pins pushed down
C44611-R22AV	992842		Output (9mV - 13mV) erratic Failed in Q.A. Did not reach end user
C44611-R22AV	999922		Output (9mV - 13mV) 15.4 n Move Item to location 126553 n Test Equipment Barcode 126553 n Failed in Q.A. Did not reach end user
C43690-R17AH	983902		Output (10.5mV - 13.5mV) 14.0 n Move Item to location 126541 n Test Equipment Barcode 126541 n Failed in Q.A. Did not reach end user
C43690-R17AH	983904		Output (10.5mV - 13.5mV) 14.1 n Move Item to location 126541 n Test Equipment Barcode 126541 n Output (10.5mV - 13.5mV) 14.1 n Move Item to location 126541 n Test Equipment Barcode 126541 n Failed in Q.A. Did not reach end user
C43690-R17AH	983907		Output (10.5mV - 13.5mV) 13.9 n Move Item to location 126541 n Test Equipment Barcode 126541 n Failed in Q.A. Did not reach end user
C43690-R17AH	983915		Output (10.5mV - 13.5mV) 14.7 n Move Item to location 126541 n Test Equipment Barcode 126541 n Failed in Q.A. Did not reach end user
C43690-R17AH	983918		Output (10.5mV - 13.5mV) 14.2 n Move Item to location 126541 n Test Equipment Barcode 126541 n Output (10.5mV - 13.5mV) 14.2 n Move Item to location 126541 n Test Equipment Barcode 126541 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A952458		Output (10.5 - 13.5 mV) 17.7 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A966260		Output (10.5 - 13.5 mV) 14.2 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996136		
C44611-R22AHJR	A996137		Output (10.5 - 13.5 mV) 10.4 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996139		Output (10.5 - 13.5 mV) 9.8 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996143		Output (10.5 - 13.5 mV) 00.1 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996145		Output (10.5 - 13.5 mV) 9.8 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996147		Failed 10.2mV failed in QA did not reach end user
C44611-R22AHJR	A996156		Output (10.5 - 13.5 mV) 10.4 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996156		Output (10.5 - 13.5 mV) 10.4 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996160		Output (10.5 - 13.5 mV) 10.0 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996173		Output (10.5 - 13.5 mV) 10.1 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996176		Output (10.5 - 13.5 mV) 10.3 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user

[illegible]

Part Number	Serial Number	RMA	Fault Description
C44611-R22AHJR	A996308		Output 10.3mV failed in QA. Dis not reach end user
C44611-R22AHJR	A996309		Output 10mV failed in QA did not reach customer
C44611-R22AHJR	A996317		Output (10.5 - 13.5 mV) 9.9 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996326		Out put 10.4mV
C44611-R22AHJR	A996339		Output 10.2mV Failed in QA. Did not reach end user
C44611-R22AHJR	A996345		Output 10.2mV Failed in QA Dis not reach End User
C44611-R22AHJR	A996346		Output 10.1mV failed in QA. Dis not reach end user
C44611-R22AHJR	A996349		Output (10.5 - 13.5 mV) 9.9 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996351		Output 10.4mv Failed in QA. Did not reach end user
C44611-R22AHJR	A996353		Output 10.3mV
C44611-R22AHJR	A996354		Output (10.5 - 13.5 mV) 9.8 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996355		Output 10.2mv Failed in QA did not reach end user
C44611-R22AHJR	A996357		
C44611-R22AHJR	A996358		Output (10.5 - 13.5 mV) 10.0 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996365		Output (10.5 - 13.5 mV) 10.3 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996367		Output (10.5 - 13.5 mV) 10.1 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996368		Output (10.5 - 13.5 mV) 10.3 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996371		Output (10.5 - 13.5 mV) 10.4 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996375		Output 10.4mV Failed in QA did not reach end user
C44611-R22AHJR	A996380		Out put 10.3mV failed in QA. Did not reach end user
C44611-R22AHJR	A996384		Output (10.5 - 13.5 mV) 9.9 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996386		Output (10.5 - 13.5 mV) 9.7 n 126550 n Test Equipment Barcode 102096 n Output (10.5 - 13.5 mV) 9.7 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996391		Output (10.5 - 13.5 mV) 10.4 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996394		Output (10.5 - 13.5 mV) 10.2 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996406		Output (10.5 - 13.5 mV) 10.4 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996409		Output (10.5 - 13.5 mV) 10.2 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996425		Output (10.5 - 13.5 mV) 10.1 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user