

Viamed Returns Shipping Invoice

Delivery Address:

Honeywell Healthcare Solutions GmbH RMA:

Andre Schliemann

Alter Holzhafen 18

Wismar

Germany

D-23966

Eori DE3174778

0049 3841 360204

Date: 25 Oct 2023

Viamed RN: RTS1100944

Shipper Address:

Viamed

15 Station Road

Cross Hills

Keighley

West Yorkshire

BD20 7DT

+44 (0) 1535 634542

+44 (0) 1535 635582

info@viamed.co.uk

Returns Under Warranty Value for Customs : 200 euroSee Attached Listing for Serial numbers and Faults

Part Number

0111277

P/N / Value

0111277

Tariff Code: 90181990-00

Description

Oxygen monitor MySign O Kit

Quantity

2

Part Number	Serial Number	RMA	Fault Description
0111277	103558		The unit fails to power up. There is no indication that the unit is charging when connected to the charger. The charger has been tested. Supplied with power supply Country of Origin: Germany Customs Value: 100 Euro
0111277	103561		The unit fails to power up. There is no indication that the unit is charging when connected to the charger. The charger has been tested. Supplied with power supply Country of Origin: Germany Customs value: 100 euro