

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 27 May 2025
RMA:
Viamed RN: RTS1100945

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17A	0110117 (\$27.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-17A	1
C44611-R22AV	0110132 (\$18.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	5
C74069-R33S1	0110215 Tariff Code: 90271010	R-33S1 Oxygen Sensor	1
C44611-R22A	0110122 (\$20.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-22A	1

Part Number	Serial Number	RMA	Fault Description
C43690-R17A	778271		Non Linear
C44611-R22A	919838		Low Output
C44611-R22AV	890926		Zero offset error
C44611-R22AV	890949		Zero offset error
C44611-R22AV	890971		Zero offset error
C44611-R22AV	931812		Zero offset error
C44611-R22AV	931826		Unstable output signal
C74069-R33S1	928936		High Output