## Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1017

0016269341589

Date: 31 Jan 2025 Shipper Address:

RMA: Viamed

Viamed RN: RTS1100956 15 Station Road Cross Hills

Keighley

West Yorkshire

**BD20 7DT** 

+44 (0) 1535 634542 +44 (0) 1535 635582 info@viamed.co.uk

Part Number	P/N / Value	Description	Quantity	
C43690-R17AH		•	Quartity 4	
C43090-K17AN	Tariff Code: HTS Code 9	Automotive Oxygen Sensor R-17A	4	
		027.10.2000		
O 40000 D 474	Country of Origin: USA	A 4	4	
C43690-R17A	` ,	Automotive Oxygen Sensor R-17A	4	
	Tariff Code: HTS Code 9	027.10.2000		
	Country of Origin: USA			
C44611-R22A	0110122 (\$20.00)	Automotive Oxygen Sensor R-22A	1	
	Tariff Code: HTS Code 9	027.10.2000		
	Country of Origin: USA			
C44611-R22AHJR	0110361 (\$27.00)	Oxygen sensor R-22AHJR Jikco	1	
	Tariff Code: HTS Code 9	027.10.2000		
	Country of Origin: USA			
C44611-R22AV	, .	Automotive Oxygen Sensor R-22A	2	
	Tariff Code: HTS Code 9			
	Country of Origin: USA			
C43690-R17AV	, .	Automotive Oxygen Sensor R-17A	1	
	Tariff Code: HTS Code 9		-	
	Country of Origin: USA			
	Country of Origin Cort			

Part Number	Serial Number	RMA	Fault Description
C43690-R17A	961970		Output ( 7mV - 13mV ) 5.4 mV Failed in Q.A. Did not reach end user
C43690-R17A	961975		Output ( 7mV - 13mV ) 00.1 mV Failed in Q.A. Did not reach end user
C43690-R17A	961981		Output ( 7mV - 13mV ) 00.1mV Failed in Q.A. Did not reach end user
C43690-R17A	961986		Output ( 7mV - 13mV ) 00.1mV Failed in Q.A. Did not reach end user
C44611-R22A	115184		Output ( 7mV - 13mV ) 14.1mV Failed in Q.A. Did not reach end user
C43690-R17AV	101045		Output ( 9mV - 13mV ) 8.6mv Failed in Q.A. Did not reach end user
C44611-R22AV	100276		Output ( 9mV - 13mV ) 1.6mV Failed in Q.A. Did not reach end user
C44611-R22AV	101828		Output ( 9mV - 13mV ) 8.3mV Failed in Q.A. Did not reach end user
C43690-R17AH	113526		Output ( 10.5mV - 13.5mV ) 14.5mv Failed in Q.A. Did not reach end user
C43690-R17AH	113533		Output ( 10.5mV - 13.5mV ) 14.0mV Failed in Q.A. Did not reach end user
C43690-R17AH	113541		Output ( 10.5mV - 13.5mV ) 13.9mV Failed in Q.A. Did not reach end user
C43690-R17AH	113550		Output ( 10.5mV - 13.5mV ) 14.5mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996127		Output ( 10.5 - 13.5 mV ) 10.8mV failed in QA, did not reach end user