

# Viamed Returns Shipping Invoice

Delivery Address:  
Teledyne Analytical Instruments  
Margie Read  
16830 Chestnut Street  
City Of Industry  
California  
91748-1017  
0016269341589

Date: 27 May 2025  
RMA:  
Viamed RN: RTS1100958

Shipper Address:  
Viamed  
15 Station Road  
Cross Hills  
Keighley  
West Yorkshire  
BD20 7DT  
+44 (0) 1535 634542  
+44 (0) 1535 635582  
info@viamed.co.uk

---

Part Number	P/N / Value	Description	Quantity
A37016-T1	0110051 (\$30.00 Ea) Tariff Code: 9027.10.2000 Country of Origin: USA	Teledyne Oxygen sensor T-1	2

Part Number	Serial Number	RMA	Fault Description
A37016-T1	997948		Leaking Failed in Q.A. Did not reach end user
A37016-T1	997949		Leaking Failed in Q.A. Did not reach end user