

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 23 Dec 2024
RMA:
Viamed RN: RTS1100963

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Part Number	P/N / Value	Description	Quantity
C44611-R22AHJR	0110361 (\$27.00)	Oxygen sensor R-22AHJR Jikco	22
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		

Part Number	Serial Number	RMA	Fault Description
C44611-R22AHJR	A134747		Output (10.5 - 13.5 mV) 14.6 n 126550 n Test Equipment Barcode 102096 n Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134752		Output (10.5 - 13.5 mV) 10.2 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134757		Output (10.5 - 13.5 mV) 9.9mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134771		Output (10.5 - 13.5 mV) 14.0mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134772		Output (10.5 - 13.5 mV) 10.3mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134776		Output (10.5 - 13.5 mV) 14.1mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134777		Output (10.5 - 13.5 mV) 10.0mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134778		Output (10.5 - 13.5 mV) 10.1mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134780		Output (10.5 - 13.5 mV) 10.0 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134784		Output (10.5 - 13.5 mV) 13.6 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134786		Output (10.5 - 13.5 mV) 10.0 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134790		Output (10.5 - 13.5 mV) 9.8 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A134791		Output (10.5 - 13.5 mV) 9.7mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996446		Output (10.5 - 13.5 mV) 9.5 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996447		Output (10.5 - 13.5 mV) 10.4 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996456		Output (10.5 - 13.5 mV) 9.7 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996457		Output (10.5 - 13.5 mV) 9.8mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996462		Output (10.5 - 13.5 mV) 10.2 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996464		Output (10.5 - 13.5 mV) 10.2mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996468		Output (10.5 - 13.5 mV) 10.4mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996471		Output (10.5 - 13.5 mV) 9.9 mV Failed in Q.A. Did not reach end user
C44611-R22AHJR	A996477		Output (10.5 - 13.5 mV) 10.3 mV Failed in Q.A. Did not reach end user