

Viamed Returns Shipping Invoice

Delivery Address:

Honeywell Healthcare Solutions GmbH RMA: 300273078

Andre Schliemann

Alter Holzhafen 18

Wismar

Germany

D-23966

Eori DE3174778

0049 3841 360204

Date: 05 Mar 2024

Viamed RN: RTS1100966

Shipper Address:

Viamed

15 Station Road

Cross Hills

Keighley

West Yorkshire

BD20 7DT

+44 (0) 1535 634542

+44 (0) 1535 635582

info@viamed.co.uk

Returns Under Warranty Value for Customs : 127.81 euro

See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
0110020	0110020	E1002724 OOM102 Oxygen Sensor	4
	Tariff Code: 9019200000		
E46-00-0021	0111275	Oxygen analyser OxiQuant S	1
	Tariff Code: 9018199000		

Part Number	Serial Number	RMA	Fault Description
0110020	V106469	300273078	Low Output Country of Origin: Germany Customs Value: 5.00 euro
0110020	V106470	300273078	Low output Country of Origin: Germany Customs Value: 5.00 euro
0110020	V106684	300273078	Low Out put Country of Origin: Germany Customs Value: 5.00 euro
0110020	V106685	300273078	Low Output Country of Origin: Germany Customs Value: 5.00 euro
E46-00-0021	A106340	300273078	High Reading (includes sensor s/n A111129) Reading 75% in air and wont calibrate any lower Country of origin: Germany Customs value: 107.81 euro