

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 23 Dec 2024
RMA:
Viamed RN: RTS1100972

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

| Part Number | P/N / Value | Description | Quantity |
|----------------|---|--------------------------------|----------|
| C44611-R22AHJR | 0110361 (\$27.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA | Oxygen sensor R-22AHJR Jikco | 21 |
| C44611-R22A | 0110122 (\$20.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA | Automotive Oxygen Sensor R-22A | 1 |
| C43690-R17A | 0110117 (\$27.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA | Automotive Oxygen Sensor R-17A | 2 |

| Part Number | Serial Number | RMA | Fault Description |
|----------------|---------------|-----|---|
| C43690-R17A | 962025 | | Output (7mV - 13mV) 00.1 mV Failed in Q.A. Did not reach end user |
| C43690-R17A | 962042 | | Output (7mV - 13mV) 00.2mV Failed in Q.A. Did not reach end user |
| C44611-R22A | 136741 | | Output (7mV - 13mV) 00.1mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155494 | | Output (10.5 - 13.5 mV) 10.4 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155497 | | Output (10.5 - 13.5 mV) 10.4 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155506 | | Output (10.5 - 13.5 mV) 9.7 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155528 | | Output (10.5 - 13.5 mV) 10.3 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155553 | | Output (10.5 - 13.5 mV) 9.4 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155567 | | Output (10.5 - 13.5 mV) 10.0 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155585 | | Output (10.5 - 13.5 mV) 10.2 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155593 | | Output (10.5 - 13.5 mV) 10.1 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155605 | | Output (10.5 - 13.5 mV) 10.4 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155614 | | Output (10.5 - 13.5 mV) 10.1 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155621 | | Output (10.5 - 13.5 mV) 10.1 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155629 | | Output (10.5 - 13.5 mV) 8.8 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155633 | | Output (10.5 - 13.5 mV) 10.1 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155656 | | Output (10.5 - 13.5 mV) 10.3 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155657 | | Output (10.5 - 13.5 mV) 10.3 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155663 | | Output (10.5 - 13.5 mV) 10.2 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155672 | | Output (10.5 - 13.5 mV) 10.3 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155688 | | Output (10.5 - 13.5 mV) 9.6 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155712 | | Output (10.5 - 13.5 mV) 10.2 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155715 | | Output (10.5 - 13.5 mV) 9.8 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A155716 | | Output (10.5 - 13.5 mV) 10.4 mV Failed in Q.A. Did not reach end user |