

Viamed Returns Shipping Invoice

Delivery Address:
Maxtec Inc
Keely Mitchell
2305 South 1070 West
Salt Lake City
UT 84119
Utah, USA

Date: 21 May 2024
RMA: 339288
Viamed RN: RTS1100975

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$578.61
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
R218P12-001	0111250 Tariff Code: 90181990-00	Maxtec Oxygen Analyser with in	1
R221P11-001	0111265 Tariff Code: 90181990-00	Maxtec Oxygen Analyser - Ultra	1

Part Number	Serial Number	RMA	Fault Description
R218P12-001	KB91399032	339288	Failed in QA Calibrate Button not working Country of Origin: USA Customs Value \$146.06
R221P11-001	JM20911042	339288	Cracked Screen Failed in Q.A. Did not reach end user Country of Origin: USA Customs Value: \$432.55