

# Viamed Returns Shipping Invoice

Delivery Address:  
Teledyne Analytical Instruments  
Margie Read  
16830 Chestnut Street  
City Of Industry  
California  
91748-1020  
0016269341589

Date: 23 Dec 2024  
RMA:  
Viamed RN: RTS1100980

Shipper Address:  
Viamed  
15 Station Road  
Cross Hills  
Keighley  
West Yorkshire  
BD20 7DT  
+44 (0) 1535 634542  
+44 (0) 1535 635582  
info@viamed.co.uk

---

| Part Number    | P/N / Value                        | Description                    | Quantity |
|----------------|------------------------------------|--------------------------------|----------|
| C43981-R21A    | 0110121 (\$16.00)                  | Automotive Oxygen Sensor R-21A | 1        |
|                | Tariff Code: HTS Code 9027.10.2000 |                                |          |
|                | Country of Origin: USA             |                                |          |
| C44611-R22AHJR | 0110361 (\$27.00)                  | Oxygen sensor R-22AHJR Jikco   | 1        |
|                | Tariff Code: HTS Code 9027.10.2000 |                                |          |
|                | Country of Origin: USA             |                                |          |

| Part Number    | Serial Number | RMA | Fault Description                                               |
|----------------|---------------|-----|-----------------------------------------------------------------|
| C43981-R21A    | 165155        |     | Output 9mV - 14mV 14.2 mV Failed in Q.A. Did not reach end user |
| C44611-R22AHJR | A996265       |     | Output ( 10.5 - 13.5 mV ) 9.8 mV. Did not reach end user        |