

Viamed Returns Shipping Invoice

Delivery Address:	Date: 18 Oct 2024	Shipper Address:
Teledyne Analytical Instruments	RMA:	Viamed
Margie Read	Viamed RN: RTS1100981	15 Station Road
16830 Chestnut Street		Cross Hills
City Of Industry		Keighley
California		West Yorkshire
91748-1020		BD20 7DT
0016269341589		+44 (0) 1535 634542
		+44 (0) 1535 635582
		info@viamed.co.uk

Part Number	P/N / Value	Description	Quantity
C44611-R22A	0110122 (\$20.00)	Automotive Oxygen Sensor R-22A	3
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		

Part Number	Serial Number	RMA	Fault Description
C44611-R22A	969786		Non Linear
C44611-R22A	970696		Unstable output signal
C44611-R22A	984857		Non Linear