

# Viamed Returns Shipping Invoice

Delivery Address:  
Teledyne Analytical Instruments  
Margie Read  
16830 Chestnut Street  
City Of Industry  
California  
91748-1020  
0016269341589

Date: 08 Nov 2024  
RMA:  
Viamed RN: RTS1100982

Shipper Address:  
Viamed  
15 Station Road  
Cross Hills  
Keighley  
West Yorkshire  
BD20 7DT  
+44 (0) 1535 634542  
+44 (0) 1535 635582  
info@viamed.co.uk

---

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$27.00 Ea) Tariff Code: 9027.10.2000 Country of Origin: USA	Teledyne Oxygen sensor R-17MED	2
C74069-R33S	0110214 Tariff Code: 90271010	R-33S Oxygen Sensor	2

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	164924		Output ( 7mV - 13mV ) 5.4mV Failed in Q.A. Did not reach end user
C43690-R17MED	164957		Output ( 7mV - 13mV ) 00.mV Failed in Q.A. Did not reach end user
C74069-R33S	972264		High Output Failed in Q.A. Did not reach end user
C74069-R33S	972269		High Output Failed in Q.A. Did not reach end user