

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Cynar Magno
16830 Chestnut Street
City Of Industry
California
91748-1017
01626-934-1556

Date: 05 Sep 2024
RMA: 0060628
Viamed RN: RTS1100992

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$987
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17AH	0110137 (\$33.00)	Automotive Oxygen Sensor R-17A	25
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
C43690-R17A	0110117 (\$27.00)	Automotive Oxygen Sensor R-17A	6
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		

Part Number	Serial Number	RMA	Fault Description
C43690-R17A	105217	0060628	Output (7mV - 13mV) 00.2 mV Failed in Q.A. Did not reach end user
C43690-R17A	105222	0060628	Output (7mV - 13mV) LEAKING Failed in Q.A. Did not reach end user
C43690-R17A	962084	0060628	Output (7mV - 13mV) 00.1 mV Failed in Q.A. Did not reach end user
C43690-R17A	962088	0060628	Output (7mV - 13mV) 00.1 mV Failed in Q.A. Did not reach end user
C43690-R17A	962093	0060628	Output (7mV - 13mV) 00.1 n mV Failed in Q.A. Did not reach end user
C43690-R17A	965049	0060628	Output (7mV - 13mV) 5.5 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176074	0060628	Output (10.5mV - 13.5mV) 13.66 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176078	0060628	Output (10.5mV - 13.5mV) 13.67 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176082	0060628	Output (10.5mV - 13.5mV) 13.57 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176088	0060628	Output (10.5mV - 13.5mV) 13.76 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176090	0060628	Output (10.5mV - 13.5mV) 13.57 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176092	0060628	Output (10.5mV - 13.5mV) 13.57 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176096	0060628	Output (10.5mV - 13.5mV) 13.96 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176100	0060628	Output (10.5mV - 13.5mV) 13.68 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176101	0060628	Output (10.5mV - 13.5mV) 13.76 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176120	0060628	Output (10.5mV - 13.5mV) 13.57 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176123	0060628	Output (10.5mV - 13.5mV) 14.75 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176128	0060628	Output (10.5mV - 13.5mV) 13.68 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176134	0060628	Output (10.5mV - 13.5mV) 13.57 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176138	0060628	Output (10.5mV - 13.5mV) 14.06 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176146	0060628	Output (10.5mV - 13.5mV) 14.06 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176148	0060628	Output (10.5mV - 13.5mV) 13.68 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176154	0060628	Output (10.5mV - 13.5mV) 14.06 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176155	0060628	Output (10.5mV - 13.5mV) 13.57 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176160	0060628	Output (10.5mV - 13.5mV) 113.96mV Failed in Q.A. Did not reach end user
C43690-R17AH	176165	0060628	Output (10.5mV - 13.5mV) 13.76 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176173	0060628	Output (10.5mV - 13.5mV) 14.06 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176193	0060628	Output (10.5mV - 13.5mV) 13.57 mV Failed in Q.A. Did not reach end user

Part Number	Serial Number	RMA	Fault Description
C43690-R17AH	176197	0060628	Output (10.5mV - 13.5mV) 13.86 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176199	0060628	Output (10.5mV - 13.5mV) 13.86 mV Failed in Q.A. Did not reach end user
C43690-R17AH	176204	0060628	Output (10.5mV - 13.5mV) 14.06 mV Failed in Q.A. Did not reach end user