

VST Returns Shipping Invoice

Delivery Address:

Honeywell Healthcare Solutions GmbH RMA: 300290069

Karen Bull

Alter Holzhafen 18

Wismar

Germany

D-23966

Eori DE3174778

0049 3841 360204

Date: 02 Oct 2025

VST RN: RTS31001050

Shipper Address:

VST

15 Station Road

Cross Hills

Keighley

West Yorkshire

BD20 7DT

+44 (0) 1535 634900

+44 (0) 1535 635582

info@vandagraphst.com

Returns Under Warranty Value for Customs : 25.00 euroSee Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
8010018	8010018 Tariff Code: 9019209000	Oxygen Sensor - T22D	1
8010014	8010014 Tariff Code: 9019209000	Oxygen Sensor - K-22D	1
8010004	8010004 Tariff Code: 9019209000	Oxygen Sensor - R17JJ-CCR	3

Part Number	Serial Number	RMA	Fault Description
8010004	JJ157582	300290069	Electrolyte Leakage Country of Origin: Germany Customs Value 5.00 Euro each
8010004	JJ157590	300290069	Electrolyte Leakage Country of Origin: Germany Customs Value 5.00 Euro each
8010004	JJ157593	300290069	Electrolyte Leakage Country of Origin: Germany Customs Value 5.00 Euro each
8010014	K104861	300290069	No Output Country of Origin: Germany Customs Value 5.00 Euro each
8010018	A106177	300290069	No Output Country of Origin: Germany Customs Value 5.00 Euro each

Part Number	Serial No.	RMA	Fault Report
8010004	JJ157582	300290069	
8010004	JJ157590	300290069	
8010004	JJ157593	300290069	
8010014	K104861	300290069	
8010018	A106177	300290069	