

VST Returns Shipping Invoice

Delivery Address:

Honeywell Healthcare Solutions GmbH RMA:

Karen Bull

Alter Holzhafen 18

Wismar

Germany

D-23966

Eori DE3174778

0049 3841 360204

Date: 14 Jan 2026

VST RN: RTS31001062

Shipper Address:

VST

15 Station Road

Cross Hills

Keighley

West Yorkshire

BD20 7DT

+44 (0) 1535 634900

+44 (0) 1535 635582

info@vandagraphst.com

Returns Under Warranty Value for Customs : 50.00 Euro

See Attached Listing for Serial numbers and Faults

Part Number

8010004

P/N / Value

8010004

Tariff Code: 9019209000

Description

Oxygen Sensor - R17JJ-CCR

Quantity

10

Part Number	Serial Number	RMA	Fault Description
8010004	JJ157979		Unstable output underwater.
8010004	JJ158583		Unstable output underwater.
8010004	JJ158587		Unstable output underwater.
8010004	JJ158591		Unstable output underwater.
8010004	JJ158593		Unstable output underwater.
8010004	JJ159007		Low Output
8010004	JJ161591		Unstable output underwater.
8010004	JJ162611		Unstable output underwater.
8010004	JJ163355		Unstable output underwater.
8010004	JJ163987		Unstable output underwater.

Part Number	Serial No.	RMA	Fault Report
8010004	JJ157979		
8010004	JJ158583		
8010004	JJ158587		
8010004	JJ158591		
8010004	JJ158593		
8010004	JJ159007		
8010004	JJ161591		
8010004	JJ162611		
8010004	JJ163355		
8010004	JJ163987		