

VST Returns Shipping Invoice

Delivery Address:

Honeywell Healthcare Solutions GmbH RMA: 3002749610

Andre Schliemann

Alter Holzhafen 18

Wismar

Germany

D-23966

Eori DE3174778

0049 3841 360204

Date: 21 Sep 2024

VST RN: RTS3100993

Shipper Address:

VST

15 Station Road

Cross Hills

Keighley

West Yorkshire

BD20 7DT

+44 (0) 1535 634900

+44 (0) 1535 635582

info@vandagraphst.com

Returns Under Warranty Value for Customs : 30.00 euro

See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
8010004	8010004 Tariff Code: 9019209000	Oxygen Sensor - R17JJ-CCR	6

Part Number	Serial Number	RMA	Fault Description
8010004	JJ146531	3002749610	Unstable output
8010004	JJ147973	3002749610	Unstable output
8010004	JJ151520	3002749610	No Output
8010004	JJ151521	3002749610	Unstable output
8010004	JJ152213	3002749610	Unstable output
8010004	JJ152830	3002749610	No Output