

Viamed Returns Shipping Invoice

Delivery Address:	Date: 23 Oct 2025	Shipper Address:
Bluepoint Medical GmbH & Co. KG (Europe)	IRMA:	Viamed
Anja Merz	Viamed RN: RTS11001053	15 Station Road
An der Trave 15		Cross Hills
Selmsdorf		Keighley
Germany		West Yorkshire
23923		BD20 7DT
0049 38823 5488020		+44 (0) 1535 634542
		+44 (0) 1535 635582
		info@viamed.co.uk

Part Number	P/N / Value	Description	Quantity
13090532002	4420507	IRMA CO2 analyser Includes:	1
	Tariff Code: 90181990-00		

Part Number	Serial Number	RMA	Fault Description
13090532002	IR00134218		The check CO2 adaptor message is displayed. Tried new lemo lead, still displays message.

Part Number
13090532002

Serial No.
IR00134218

RMA

Fault Report