

EC Declaration of Conformity

We declare with sole responsibility that following product comply with the basic requirements according the Council Directive 93/42/EEC in the current valid version (Annex I) concerning medical devices.

Product: **Oxygen Sensor**

Typ:

OOM101	OOM104	OOM110	OOM201
OOM102	OOM105	OOM111	OOM202
OOM102-1	OOM106	OOM112	OOM202-1
OOM103	OOM107		OOM202-2
OOM103-1	OOM107-2		OOM202-2S
OOM103-1M			OOM204

Classification: **Class IIa**
(RL 93/42/EEC, Annex IX)

CE marking: 

Notified Body: TÜV SÜD Product Service GmbH, Munic, Germany

Conformity Assessment Process: Annex II, section 3 of the Directive 93/42/EEC

Particular product standard applied: EN ISO 21647:2009

Issued by: **ENVITEC-WISMAR GMBH**
Alter Holzhafen 18
D-23966 Wismar
Germany

Valid from: **2012-10-01**

Place, Date: **Wismar, 2012-10-01**

Authorized Signature:



Marcus Ostländer
Managing Director