

DVM Calibration record

1	DVM Calibration Index Number	CE0179
2	DVM Bar code	682836
3	Equipment Used	MICROCAL 1030
4	Serial number Test Equipment	6839K6
5	Calibration Certificate Number	1512404912
6	Test Equipment Calibration Index Number	CE076
	Micro Cal @	DVM Reading
7	0mV +/- 0.1	0.055mV
8	5mV +/- 0.1	5.031mV
9	10mV +/- 0.1	10.049mV
10	15mV +/- 0.1	15.086mV
11	20mV +/- 0.1	20.117mV
12	25mV +/- 0.1	25.152mV
13	30mV +/- 0.1	30.142mV
14	35mV +/- 0.1	35.153mV
15	40mv +/- 0.1	40.130mV
16	Pass / Fail	PASS
17	Tested By	M. GREEN
18	Tested On	4-12-17
19	Signed	M. Green.

1. Enter the details of the unit being tested in boxes 1 – 2
2. Enter the details of the calibrated Micro Cal unit being used to test against the equipment. In to boxes 3 -6

Ensure the calibrated unit is still inside its calibration period before you begin.
3. set the micro cal to the required settings (7 – 15) and enter the readings on the DVM in the Boxes
4. Assuming the units passes the tests, date and sign the form (17 – 19). Scan the form and upload to the issue requesting calibration and complete the Issue.
 If the unit fails the test return the Unit to RS for calibration