

# EU DECLARATION OF CONFORMITY

---

**APPLICATION OF COUNCIL DIRECTIVE** : Annex II, Section 3.2, of MDD 93/42/EEC, as amended by Directive 2007/47/EC RoHS Directive 2011/65/EU, along with Annex IV, item 1(b) exception

**STANDARD APPLIED** : EN 60601-1, 3<sup>rd</sup> Edition  
EN 60601-1-2, EN60601-1-6, EN 14971:2012,  
EN ISO 80601-2-55:2011

**MANUFACTURER'S NAME** : TELEDYNE ANALYTICAL INSTRUMENTS


**MANUFACTURER'S ADDRESS** : 16830 Chestnut Street  
City of Industry, CA 91748  
U.S.A.

**TYPE OF EQUIPMENT** : Oxygen Sensor

**MODEL NUMBER** : R17-MEDV

---

**I, THE UNDERSIGNED, HEREBY DECLARE THAT THE EQUIPMENT SPECIFIED ABOVE CONFORMS TO THE ABOVE DIRECTIVE**

**SIGNATURE:** 

**FULL NAME:** Angel Alegria

**POSITION:** New Products Manager

**Date:** 12-13-17

**PLACE:** City of Industry, California