

CERTIFICATE OF CALIBRATION

Issued by: RS Components Ltd

Date Issued: 27 Feb 2018

Certificate No.

1517284



RS Calibration

Calibration and Repair Service

Page 1 of 2 Pages

DPN 175, Lamas Rd,
Weldon Industrial Est
Corby, Northants, NN17 9RS

Tel: 01536 405545
Fax: 01536 401590


Simon Perkins

Client	VIAMED LTD KEIGHLEY WEST YORKSHIRE BD20 7DT
Instrument	Digitron P200L i.s. 0 to 500 mbar Digital Manometer
Serial No.	10198
Client Reference	CE091 102075
Procedure ID.	188_1175_500mbar Rev. P3
Date of Calibration	27 Feb 2018

Equipment Used to Carry Out Calibration

Ruska Deadweight Tester
Weights

Equipment ID.

Cal 711
Cal 713, 714

The measurements reported in this certificate were carried out using equipment whose values are traceable to national standards.

All procedures employed and results reported are in compliance with the requirements of the International Standard ISO/IEC 17025:2005.

The management controls of the RS Calibration Laboratory are registered under the British Standard BS EN ISO 9001 : 2008 No. RS 00362.

Uncertainties

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%.

This certificate reports recorded values for the instrument 'As Received'.



RS Components takes its environmental responsibilities very seriously and as such has printed this double sided document in black and white, on paper from sustainable sources.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

Certificate No.
1517284



Calibration and Repair Service

Page 2 of 2 Pages

Environment

The ambient temperature and relative humidity throughout the calibration were $(20 \pm 2) ^\circ\text{C}$ and $(40 \pm 20) \% \text{RH}$ respectively.

Method

The instrument was calibrated using a dead weight pressure standard, applying a set mass to achieve nominal values of gauge pressure and recording the instrument's displayed values.

Prior to calibration the instrument was exercised to full scale pressure and then returned to zero.

The pressure medium used was nitrogen.

Range Gauge	Applied Value	Instrument Indicated	Percentage of Tolerance	L.S.D. Stability
Lo	0 mbar	0.0 mbar	0 %	0
	100 mbar	99.8 mbar	20 %	0
	190 mbar	190.0 mbar	0 %	0
Hi	300 mbar	300 mbar	0 %	0
	400 mbar	400 mbar	0 %	0
	500 mbar	500 mbar	0 %	1

Measurement uncertainty of the above recorded values: $\pm (0.01\% + 1 \text{ L.S.D.})$

END OF CALIBRATION

CALIBRATED BY:- SPP

Compliance to Specification

The specification published by the manufacturer and found in the instrument's handbook has been used to determine performance at the measured points.

Reported values

The uncertainties quoted refer to the applied values, which include any identified contribution of the instrument under test and not to the ability of the instrument to maintain its calibration.