



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.

Website: www.viamed.co.uk. Email: info@viamed.co.uk.

Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	046019
Model.	TT 480.	Time & Date of Service / Calibration.	PT 7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.		P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	41.4	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.2	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH ₂ O +/- 1 (28.4 – 30.4 mBar)	30.0	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label. Serial No. Label. Tom Thumb Label	✓ ✓ ✓	P

Service.	Parts Replaced.
Changed "O" Rings	✓ Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓ Adjustable Valve
Re-calibrated Tom Thumb	✓ Adjustable Valve Internal Parts
	Safety Valve
	Safety Valve Internal Parts
	Pressure gauge
	Pressure Gauge Face
	"O" Rings
	Outlet
	Inlet
	Blanking Bolts
	Rail Clamp
	Main Body Block

Calibration: S. HARDAKER Signed: BS

**VIAMED Ltd.**

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
 Website: www.viamed.co.uk. Email: info@viamed.co.uk.
 Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040076
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/T
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: (@ 10 Lpm)	Minimum: 8 cmH ₂ O.	✓	p
Manometer	Adjustable Valve: (@ 10 Lpm.)	Tolerance: 42.1 +/- 1 mBar	42.0	p
Manometer	Precision Valve: (@ 10 Lpm.)	44.1 mBar +/- 1	44.1	p
Manometer	Pressure Gauge Test: (@ 30 cmH ₂ O.)	Tolerance: 30cmH ₂ O +/- 1 (28.4 – 30.4 mBar)	30.0	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: S. HARDAKER Signed: MM



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.

Website: www.viamed.co.uk. Email: info@viamed.co.uk.

Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040028
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	41.4	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	43.4	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- (28.4 – 30.4 mBar)	30.4	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: S. HARDAKER Signed:

**VIAMED Ltd.**

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.

Website: www.viamed.co.uk. Email: info@viamed.co.uk.

Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	060025
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	42.4	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.1	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- (28.4 – 30.4 mBar)	29.6	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

S. HARDAKER
 Calibration: Signed: 



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
Website: www.viamed.co.uk. Email: info@viamed.co.uk.
Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040021
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	42.7	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.8	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH ₂ O +/- 1 (28.4 – 30.4 mBar)	29.9	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: S. HARDAKER Signed: BMM



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
Website: www.viamed.co.uk. Email: info@viamed.co.uk.
Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	5401379
Model.	TT 480.	Time & Date of Service / Calibration.	12/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.		p
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar		p
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1		p
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH ₂ O +/- 1 (28.4 – 30.4 mBar)		P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label. Serial No. Label. Tom Thumb Label	✓ ✓ ✓	P

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: D. LAMB..... Signed: P. East.....

**VIAMED Ltd.**

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
 Website: www.viamed.co.uk. Email: info@viamed.co.uk.
 Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040020
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	41.8	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	45.0	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH ₂ O +/- 1 (28.4 – 30.4 mBar)	29.6	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.
Changed "O" Rings	✓ Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓ Adjustable Valve
Re-calibrated Tom Thumb	✓ Adjustable Valve Internal Parts Safety Valve Safety Valve Internal Parts Pressure gauge Pressure Gauge Face "O" Rings
	✓ Outlet Inlet Blanking Bolts Rail Clamp Main Body Block

Calibration: Signed: 



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.

Website: www.viamed.co.uk. Email: info@viamed.co.uk.

Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	VMC9069
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/19

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	41.5	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	43.7	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH ₂ O +/- 1 (28.4 – 30.4 mBar)	28.9	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.
Changed "O" Rings	✓ Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓ Adjustable Valve
Re-calibrated Tom Thumb	✓ Adjustable Valve Internal Parts
	Safety Valve
	Safety Valve Internal Parts
	Pressure gauge
	Pressure Gauge Face
	"O" Rings
	Outlet
	Inlet
	Blanking Bolts
	Rail Clamp
	Main Body Block

Calibration: S HARDAKER Signed:

Supplier - Subcontractor Assessment Form

Ref: Form 14 Issue: 2 Page: 1 of 1



Supplier / Sub Contractor to Complete				
Full Company Name and Postal Address		Contact Name (General)		
		Contact Name QHSE		
		Telephone Number		
Post Code		E Mail Address		
Services provided by your company are: Please give a brief description of the products and services provided				
Trade Accreditation / Operating Licences/ Insurances certificates (Please attach a copies of Licences, Liability insurances, Trade certificates)				
Certificate type	Issued by	Certificate No.	Issue / Expiry Date	
Public Liability Certificate				
Product Liability Certificate				
QHSE MANAGEMENT SYSTEMS (PLEASE INDICATE YOUR STATUS ON EACH CATEGORY)				
Documented Quality Management System? <input type="checkbox"/> ISO9001 Accredited system <input type="checkbox"/> In progress with timescale (Please specify expected date) <input type="checkbox"/> Only Policy Statement (Please provide a copy) <input type="checkbox"/> None (Please explain why)		Comments		
Documented Health and Safety Management System? <input type="checkbox"/> OHSAS 18001 Accredited system <input type="checkbox"/> In progress with timescale (Please specify expected date) <input type="checkbox"/> Only Policy Statement (Please provide a copy) <input type="checkbox"/> None (Please explain why)				
Documented Environment Management System? <input type="checkbox"/> ISO14001 Accredited system <input type="checkbox"/> In progress with timescale (Please specify expected date) <input type="checkbox"/> Only Policy Statement (Please provide a copy) <input type="checkbox"/> None (Please explain why)		Comments		
For Office Use Only, (JFD QHSE to Complete)				
Assessment decision	<input type="checkbox"/> Accept	<input type="checkbox"/> Monitor	<input type="checkbox"/> Audit	<input type="checkbox"/> Reject
Assessed by			DATE ASSESSED	
Assessment comments / Supplier history			CAR/ NCR	Date

**VIAMED Ltd.**

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
 Website: www.viamed.co.uk. Email: info@viamed.co.uk.
 Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	VMC 9073
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	43.0	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.4	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	29.7	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.
Changed "O" Rings	✓ Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓ Adjustable Valve
Re-calibrated Tom Thumb	✓ Adjustable Valve Internal Parts Safety Valve Safety Valve Internal Parts Pressure gauge Pressure Gauge Face "O" Rings
	✓ Outlet Inlet Blanking Bolts Rail Clamp Main Body Block

Calibration: Signed: 



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
Website: www.viamed.co.uk. Email: info@viamed.co.uk.
Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	VMA 9009	
Model.	TT 480.	Time & Date of Service / Calibration.		

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	p
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	41.4	p
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.6	p
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH ₂ O +/- 1 (28.4 – 30.4 mBar)	30.1	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.
Changed "O" Rings	✓ Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓ Adjustable Valve
Re-calibrated Tom Thumb	✓ Adjustable Valve Internal Parts
	Safety Valve
	Safety Valve Internal Parts
	Pressure gauge
	Pressure Gauge Face
	"O" Rings
	Outlet
	Inlet
	Blanking Bolts
	Rail Clamp
	Main Body Block

Calibration: Signed: 

**VIAMED Ltd.**

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
 Website: www.viamed.co.uk. Email: info@viamed.co.uk.
 Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040022
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.		P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	42.4	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	45.0	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH2O +/- 1 (28.4 – 30.4 mBar)	30.3	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.
Changed "O" Rings	✓ Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓ Adjustable Valve
Re-calibrated Tom Thumb	✓ Adjustable Valve Internal Parts
	Safety Valve
	Safety Valve Internal Parts
	Pressure gauge
	Pressure Gauge Face
	"O" Rings
	Outlet
	Inlet
	Blanking Bolts
	Rail Clamp
	Main Body Block

Calibration: S. HARDAKER Signed: MM



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
Website: www.viamed.co.uk. Email: info@viamed.co.uk.
Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040086
Model.	TT 480.	Time & Date of Service / Calibration.	02/08/17 19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	p
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	42.9	p
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	43.5	p
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	30.2	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label. Serial No. Label. Tom Thumb Label	✓ ✓ ✓	P

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: S. HARDAKER Signed: 



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
Website: www.viamed.co.uk. Email: info@viamed.co.uk.
Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040085
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	42.9	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.5	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	29.8	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.
Changed "O" Rings	✓ Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓ Adjustable Valve
Re-calibrated Tom Thumb	✓ Adjustable Valve Internal Parts Safety Valve Safety Valve Internal Parts Pressure gauge Pressure Gauge Face "O" Rings
	✓ Outlet Inlet Blanking Bolts Rail Clamp Main Body Block

Calibration: S. HARDAKER Signed: 

**VIAMED Ltd.**

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
 Website: www.viamed.co.uk. Email: info@viamed.co.uk.
 Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	VMC 907C
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	42.6	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.1	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	30.4	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label. Serial No. Label. Tom Thumb Label	✓ ✓ ✓	P

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: S. HARDAKER Signed:

**VIAMED Ltd.**

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
 Website: www.viamed.co.uk. Email: info@viamed.co.uk.
 Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	VMA 9001
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	42.6	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	43.9	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	30.0	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: S. HARDAKER Signed:



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
 Website: www.viamed.co.uk. Email: info@viamed.co.uk.
 Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040024
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	41.3	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	43.4	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	29.3	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label. Serial No. Label. Tom Thumb Label	✓ ✓ ✓	P

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

S. HARDAKER
 Calibration: Signed: 



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
Website: www.viamed.co.uk. Email: info@viamed.co.uk.
Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	VMC 9072
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	p
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	42.3	p
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	43.1	p
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	29.3	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: S HARDAKER Signed:

**VIAMED Ltd.**

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
 Website: www.viamed.co.uk. Email: info@viamed.co.uk.
 Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040023
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	41.5	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	43.7	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	28.6	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: Signed: 



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
Website: www.viamed.co.uk. Email: info@viamed.co.uk.
Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.


TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	0600070
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	42.5	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.1	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	30.4	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label. Serial No. Label. Tom Thumb Label	✓ ✓ ✓	P

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: S. HARDAKER Signed: 



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.

Website: www.viamed.co.uk. Email: info@viamed.co.uk.

Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	CO2
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	41.2	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.3	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	29.1	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label. Serial No. Label. Tom Thumb Label	✓ ✓ ✓	P

Service.	Parts Replaced.
Changed "O" Rings	✓ Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓ Adjustable Valve
Re-calibrated Tom Thumb	✓ Adjustable Valve Internal Parts
	Safety Valve
	Safety Valve Internal Parts
	Pressure gauge
	Pressure Gauge Face
	"O" Rings
	Outlet
	Inlet
	Blanking Bolts
	Rail Clamp
	Main Body Block

Calibration: S HARDAKER Signed: SS



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
Website: www.viamed.co.uk. Email: info@viamed.co.uk.
Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	VM69068
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/11

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F Calibration
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	43.0	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.7	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH ₂ O +/- 1 (28.4 – 30.4 mBar)	30.1	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label.	✓	P
		Serial No. Label.	✓	
		Tom Thumb Label	✓	

Service.	Parts Replaced.
Changed "O" Rings	✓ Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓ Adjustable Valve
Re-calibrated Tom Thumb	✓ Adjustable Valve Internal Parts Safety Valve Safety Valve Internal Parts Pressure gauge Pressure Gauge Face "O" Rings
	✓ Outlet Inlet Blanking Bolts Rail Clamp Main Body Block

Calibration: Signed: 



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
Website: www.viamed.co.uk. Email: info@viamed.co.uk.
Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040091
Model.	TT 480.	Time & Date of Service / Calibration.	19/11/11

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	41.7	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	44.7	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH2O +/- 1 (28.4 – 30.4 mBar)	30.3	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label. Serial No. Label. Tom Thumb Label	✓ ✓ ✓	P

Service.	Parts Replaced.
Changed "O" Rings	✓ Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓ Adjustable Valve
Re-calibrated Tom Thumb	✓ Adjustable Valve Internal Parts Safety Valve Safety Valve Internal Parts Pressure gauge Pressure Gauge Face "O" Rings
	✓ Outlet
	Inlet
	Blanking Bolts
	Rail Clamp
	Main Body Block

Calibration: S.HARDAKER Signed: MM

**VIAMED Ltd.**

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.

Website: www.viamed.co.uk. Email: info@viamed.co.uk.

Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	040027
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: (@ 10 Lpm)	Minimum: 8 cmH ₂ O.	✓	p
Manometer	Adjustable Valve: (@ 10 Lpm.)	Tolerance: 42.1 +/- 1 mBar	41.3	p
Manometer	Precision Valve: (@ 10 Lpm.)	44.1 mBar +/- 1	44.9	p
Manometer	Pressure Gauge Test: (@ 30 cmH ₂ O.)	Tolerance: 30cmH20 +/- 1 (28.4 – 30.4 mBar)	29.6	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label. Serial No. Label. Tom Thumb Label	✓ ✓ ✓	P

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: S. HARDAKER Signed: MM



VIAMED Ltd.
15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.
Website: www.viamed.co.uk. Email: info@viamed.co.uk.
Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

TOM THUMB CALIBRATION / SERVICE SHEET.

Description.	Tom Thumb.	Serial No.	VMC9071
Model.	TT 480.	Time & Date of Service / Calibration.	19/7/17

Record the manometer reading in millibars below, ensuring the TT480 meets the limits specified.

Test Equipment	Test	Specification	Reading	P/F
			Calibration	
Snoop liquid.	Check all ports / connections for leaks.	No bubbling.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm	Minimum: 8 cmH ₂ O.	✓	P
Manometer	Adjustable Valve: @ 10 Lpm.	Tolerance: 42.1 +/- 1 mBar	41.7	P
Manometer	Precision Valve: @ 10 Lpm.	44.1 mBar +/- 1	43.7	P
Manometer	Pressure Gauge Test: @ 30 cmH ₂ O.	Tolerance: 30cmH20 +/-1 (28.4 – 30.4 mBar)	29.7	P
Visual check.	Gauge cannot be removed without tools.		✓	P
Visual check.	Gauge appears straight.		✓	P
Visual check.	All adjustable settings are set to a minimum.		✓	P
Visual check.	Labels are attached.	CE Label. Serial No. Label. Tom Thumb Label	✓ ✓ ✓	P

Service.	Parts Replaced.	
Changed "O" Rings	✓	Diamond Copper Washer
Re-greased Adjustable Valve Internal Parts	✓	Adjustable Valve
Re-calibrated Tom Thumb	✓	Adjustable Valve Internal Parts
		Safety Valve
		Safety Valve Internal Parts
		Pressure gauge
		Pressure Gauge Face
		"O" Rings
		Outlet
		Inlet
		Blanking Bolts
		Rail Clamp
		Main Body Block

Calibration: S. HARDAKER Signed: 