

# CERTIFICATE OF CALIBRATION

Issued by: RS Components Ltd

Date Issued: 13 Jun 2018

Certificate No. 1531195



## RS Calibration

Calibration and Repair Service

Page 1 of 3 Pages

DPN 175, Lammas Rd,  
Weldon Industrial Est  
Corby, Northants, NN17 9RS

Tel: 01536 405545  
Fax: 01536 401590

A handwritten signature in black ink.

Sharleen Forde

### Client

VIAMED LTD  
KEIGHLEY  
WEST YORKSHIRE  
BD20 7DT

### Instrument

Thurlby Thandar TF930 Frequency Counter

### Serial No.

425571

### Client Reference

N/A

### Procedure ID.

665\_4919\_TF930 Rev P8

### Date of Calibration

13 Jun 2018

### Remarks

This certificate reports recorded values for the instrument 'As Received'.

### Uncertainties

The reported expanded uncertainties are based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.



RS Components takes its environmental responsibilities very seriously and as such has printed this double sided document in black and white, on paper from sustainable sources.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0310

Certificate No.

1531195



Calibration and Repair Service

Page 2 of 3 Pages

## Environment

The ambient temperature and relative humidity throughout the calibration were  $(20 \pm 2)^\circ\text{C}$  and  $(40 \pm 20)\%$ RH respectively.

## Method

Prior to the calibration the instrument was held within the laboratory for a period of not less than 1 hour.

The instrument readings are recorded in the table below and are those with the instrument set for an appropriate gate time.

Function & Range	Applied Value	Gate Time	Nominal Amplitud	Unit Reading	UUT L.S.D Stability
<b>Input A</b>					
50 Hz	50 Hz	1 s	42 mV	50.000029 Hz	0
110 Hz	110 Hz	1 s	42 mV	110.00006 Hz	0
110 Hz	110 Hz	10 s	42 mV	110.000006 Hz	0
1.1 kHz	1.1 kHz	1 s	42 mV	1.1000006 kHz	0
1.1 kHz	1.1 kHz	10 s	42 mV	1.10000065 kHz	0
11 kHz	11 kHz	0.3 s	42 mV	11.000007 kHz	1
11 kHz	11 kHz	1 s	42 mV	11.000007 kHz	0
11 kHz	11 kHz	10 s	42 mV	11.0000065 kHz	0
110 kHz	110 kHz	1 s	42 mV	110.00006 kHz	0
1100 kHz	1100 kHz	1 s	42 mV	1100.0007 kHz	1
11 MHz	11 MHz	1 s	42 mV	11.000007 MHz	1
11 MHz	11 MHz	10 s	42 mV	11.0000065 MHz	0
110 MHz	110 MHz	1 s	42 mV	110.00007 MHz	0
<b>Input B</b>					
110 MHz	110 MHz	0.3 s	70 mV	110.00007 MHz	1
110 MHz	110 MHz	1 s	70 mV	110.00007 MHz	1
110 MHz	110 MHz	10 s	70 mV	110.000065 MHz	0
2000 MHz	2000 MHz	1 s	70 mV	2000.0012 MHz	0
<b>Period</b>					
2 MHz	2 MHz	1 s	42 mV	500.00092 us	0

## Measurement Uncertainties

10 Hz to 100 kHz	$\pm (1 \text{ in } 10^6 + 1 \text{ L.S.D})$
100 kHz to 1 MHz	$\pm (1 \text{ in } 10^7 + 1 \text{ L.S.D})$
1 MHz to 2 000 MHz	$\pm (5 \text{ in } 10^8 + 1 \text{ L.S.D})$