



Steve Nixon <steve.nixon.viamed@googlemail.com>

Speaker damage

1 message

Steve Nixon <steve.nixon@viamed.co.uk>
Reply-To: steve.nixon@viamed.co.uk
To: Dave Brough <david.brough@virgin.net>

29 June 2018 at 17:08

Hi Dave

I've just checked and the double sided gel sheets that we have been testing are all dated 12/16, so there is no apparent degradation.

We will build up several speaker assemblies and test your version pad against some of our versions.

If you look at the attached images you will see why historically we have been cautious.

- 1) The output is not as good as the bag.
- 2) The Panasonic version speaker is damaged by long term contact with similar gels.
- 3) More alarmingly any oil that migrates across or leaks through the bag materials would cause even more severe damage to the speaker. The material holding the speaker assembly would disintegrate (turn to gel) and peel away.

With your testing did you use screw spacers on both sides?

Regards

--
Steve

Steve Nixon
Director - Viamed Ltd.

Office: +44 (0)1535 634542
Mobile: +44 (0)7850 252267

2 attachments



Speaker pics 1.JPG
1071K



Speaker pics 2.JPG
1343K