



Steve Nixon <steve.nixon.viamed@googlemail.com>

Re: DFB pricing

1 message

Bob Kalish <bkalish@maxtec.com>
To: Steve Nixon <steve.nixon@viamed.co.uk>

25 July 2018 at 19:55

Steve,

It is not a silly question actually. The DFBs are dependent on the gas inlet pressure.

So just like with the MB2, there is only one version that is suitable for your 60 PSI in Europe.

It is p/n R219P77-400 and it is \$139.

And just like the MB2 you purchased, it indicates both 0-15 and 0-5 LPM, which is the most popular increments that should satisfy the vast majority of blender applications in the NICU.

Yes the elbow can be fitted either way around on all microblenders.

Thanks.

Bob KalishOEM and International Sales
Senior Advisor

P: +1 (385) 549-8002

M: +1 (801) 556-7623

MAXTEC.COM

FACEBOOK | TWITTER | NEWSLETTER



How did I do today?

On Wed, Jul 25, 2018 at 11:17 AM, Steve Nixon <office@viamed.co.uk> wrote:

Hi Bob

Can you please let me know the pricing of DFB range.

Silly question: are they suitable for use on any blender irrespective of gas input pressure?

Can the 90 degree elbow connection be fitted either way around, i.e. for left or right handed connection.

--
SteveSteve Nixon
Director - Viamed Ltd.Office: +44 (0)1535 634542
Mobile: +44 (0)7850 252267