



Showing From :01 Feb 2018

## QA Review - Failures

**Check the Previous Months Figures, First number is quantity Failed QA, Second is the number that have passed through QA, and the Final number is the Percentage of units Failing QA.**

**Raise a non Conformance report for any where the % is Greater than 5%**

Stock Reference	Description	Feb 2018	Mar 2018	Apr 2018	May 2018	Jun 2018	Jul 2018	
0110122	Oxygen sensor - Automotive R-22A	<a href="#">4 / 373</a> <a href="#">1.07%</a>	<a href="#">2 / 1003</a> <a href="#">0.20%</a>	<a href="#">0 /</a>	<a href="#">27 / 1498</a> <a href="#">1.80%</a>	<a href="#">7 / 500</a> <a href="#">1.40%</a>	<a href="#">14 / 1013</a> <a href="#">1.38%</a>	<a href="#">Raise Non Conformance review</a>
0110361	Oxygen sensor R-22AHJR Jikco OEM sensor	Examine <a href="#">3 / 114</a> <a href="#">2.63%</a>	<a href="#">1 / 80</a> <a href="#">1.25%</a>	<a href="#">0 / 39</a>	Examine <a href="#">9 / 380</a> <a href="#">2.37%</a>	<a href="#">0 /</a>	<a href="#">0 / 300</a>	<a href="#">Raise Non Conformance review</a>
0110137	Oxygen sensor - Automotive R-17AH	Examine <a href="#">10 / 132</a> <a href="#">7.58%</a>	<a href="#">0 /</a>	Examine <a href="#">11 / 97</a> <a href="#">11.34%</a>	<a href="#">0 / 66</a>	Examine <a href="#">27 / 150</a> <a href="#">18.00%</a>	<a href="#">0 /</a>	<a href="#">Raise Non Conformance review</a>
0110017	Oxygen sensor R-17MED	<a href="#">7 / 603</a> <a href="#">1.16%</a>	Examine <a href="#">2 / 55</a> <a href="#">3.64%</a>	<a href="#">6 / 966</a> <a href="#">0.62%</a>	<a href="#">0 / 1</a>	Examine <a href="#">25 / 533</a> <a href="#">4.69%</a>	<a href="#">0 / 75</a>	<a href="#">Raise Non Conformance review</a>
0110132	Oxygen sensor - Automotive R-22AV	<a href="#">2 / 498</a> <a href="#">0.40%</a>	<a href="#">15 / 1108</a> <a href="#">1.35%</a>	Examine <a href="#">13 / 409</a> <a href="#">3.18%</a>	Examine <a href="#">22 / 390</a> <a href="#">5.64%</a>	<a href="#">12 / 609</a> <a href="#">1.97%</a>	Examine <a href="#">57 / 500</a> <a href="#">11.40%</a>	<a href="#">Raise Non Conformance review</a>
0110107	Oxygen sensor - Automotive R-17A-LV	Examine <a href="#">4 / 50</a> <a href="#">8.00%</a>	<a href="#">0 / 58</a>	<a href="#">0 / 50</a>	<a href="#">0 /</a>	Examine <a href="#">4 / 100</a> <a href="#">4.00%</a>	<a href="#">0 /</a>	<a href="#">Raise Non Conformance review</a>
0110117	Oxygen sensor - Automotive R-17A	<a href="#">2 / 100</a> <a href="#">2.00%</a>	<a href="#">0 / 2</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	Examine <a href="#">7 / 80</a> <a href="#">8.75%</a>	<a href="#">Raise Non Conformance review</a>
0110127	Oxygen sensor - Automotive R-17AV	Examine <a href="#">5 / 137</a> <a href="#">3.65%</a>	Examine <a href="#">7 / 61</a> <a href="#">11.48%</a>	<a href="#">0 / 50</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">Raise Non Conformance review</a>
0110349	Oxygen sensor R21AV. HAIK OEM sensor.	<a href="#">0 /</a>	<a href="#">4 / 200</a> <a href="#">2.00%</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	Examine <a href="#">5 / 100</a> <a href="#">5.00%</a>	<a href="#">Raise Non Conformance review</a>
0110022	Oxygen sensor R-22MED	<a href="#">0 /</a>	<a href="#">0 / 10</a>	<a href="#">1 / 589</a> <a href="#">0.17%</a>	<a href="#">0 /</a>	<a href="#">0 / 301</a>	<a href="#">0 / 36</a>	<a href="#">Raise Non Conformance review</a>
0015000	Pulse oximetry Lingual Sensor	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">1 / 51</a> <a href="#">1.96%</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">Raise Non Conformance review</a>
0012101	Finger Pulse Oximeter Viamed VM 2101.	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">1 / 250</a> <a href="#">0.40%</a>	<a href="#">0 /</a>	<a href="#">0 / 250</a>	<a href="#">Raise Non Conformance review</a>
0110437	Oxygen sensor MAX-250(B)-SLE.	<a href="#">0 / 6</a>	<a href="#">0 / 10</a>	<a href="#">0 /</a>	Examine <a href="#">1 / 31</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">Raise Non Conformance</a>

					<a href="#">3.23%</a>		<a href="#">review</a>
					Examine		<a href="#">Raise Non</a>
0110001	Oxygen sensor C-1R	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">3 / 11</a>	<a href="#">0 /</a>	<a href="#">Conformance</a>
					<a href="#">27.27%</a>	<a href="#">0 /</a>	<a href="#">review</a>
0110121	Oxygen sensor - Automotive R-21A	<a href="#">0 / 8</a>	<a href="#">0 / 67</a>	<a href="#">0 / 250</a>	<a href="#">0 /</a>	<a href="#">1 / 100</a>	<a href="#">Raise Non</a>
						<a href="#">1.00%</a>	<a href="#">Conformance</a>
						<a href="#">0 / 19</a>	<a href="#">review</a>
0111230	Oxygen analyser w/o alarms - AX300-I.	<a href="#">0 /</a>	<a href="#">0 / 20</a>	<a href="#">0 / 30</a>	<a href="#">0 /</a>	Examine	<a href="#">Raise Non</a>
						<a href="#">1 / 24</a>	<a href="#">Conformance</a>
						<a href="#">4.17%</a>	<a href="#">review</a>
0111211	Oxygen analyser w/o alarms - GB300.	<a href="#">0 / 1</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	<a href="#">0 /</a>	Examine	<a href="#">Raise Non</a>
						<a href="#">2 / 2</a>	<a href="#">Conformance</a>
						<a href="#">100.00%</a>	<a href="#">review</a>