

Freigabe Label – Release for Label

☒ Label neu / new

☐ Label geändert / changed **DCR-Nr. 875**

Die Freigabe wird erteilt für – Released for:

☐ Artikelnummer – Part No.	E1002856
☐ Produkt – Product	OOD103
☒ Product Label	Sensor Label 90x15
☒ Label Primär-Verpackung /primary label	Blister Label 63,5x46,5
☐ Label Sekundär-Verpackung /secondary label	
☐ Sonstiges – Other	


Label:

 www.wittgas.com Tel +49 (0)2302 8901-0	OSGA-20 Oxygen Sensor     2020-11 S/N W123456 	 www.wittgas.com WITT-Gasetechnik GmbH & Co KG Salinger Feld 4-8 58454 Witten Deutschland/Germany OSGA-20 Oxygen Sensor  2020-11 S/N W123456     This oxygen sensor contains caustic liquid, in case of leakage avoid contact with eyes and skin. Do not puncture or staple this bag. Do not remove the oxygen sensor from this bag until ready to use. Depending upon the ambient conditions the oxygen sensor may require 30 minutes to reach signal stability after installation.
Sensor Label 90x15		Blister Label 63,5x46,5

Die Freigabe wird erteilt durch – Approved by:

	Funktion Function	Freigabe für Approval for	Unterschrift Signature	Initialen Initial	Datum Date
☒	R&D Engineer	Normenkonformität – Compliance with product standards			
☒	Produktmanager Product manager	Layout, Corporate Identity			
☒	QMB – Quality Manager	Übergeordnete Normen und Richtlinien – QMS standards & Directives			
☒	Vertrieb Sales	Kundenanforderungen – Customer requirements			
☒	FTV Production engineering	Probedruck, Stammdatenpflege Navision – Test print, maintenance of Navision material master data			
☒	Fertigungsleiter Production manager	Überführung in die Fertigung / Eintrag ins FKB – Transfer into production, entry into DHR			

Kundenfreigabe / Customer Approval

	Kunde/ Customer	Signature	Name	Date
☒	Vandagraph Sensor Technologies Ltd.		S. Nixon	25/11/20

	Created – Erstellt:	Page – Seite:
R:\PRODUKT.001\PHA\Fertigungsunterlagen\Labelfreigaben\001-27-Vandagraph-WITT-0.pdf	AO	1 / 2