



www.wittgas.com
Tel +49 (0)2302 8901-0



2020-10

S/N Wxxxxxx



OSGA-20 Oxygen Sensor



WITT-Gasetechnik GmbH & Co KG
Salinger Feld 4-8
58454 Witten
Deutschland/Germany

Tel +49 (0)2302 8901-0
Fax +49 (0)2302 8901-3
www.wittgas.com
witt@wittgas.com

OSGA-20 Oxygen Sensor



2020-10

S/N

Wxxxxxx



This oxygen sensor contains caustic liquid, in case of leakage avoid contact with eyes and skin. Do not puncture or staple this bag. Do not remove the oxygen sensor from this bag until ready to use. Depending upon the ambient conditions the oxygen sensor may require 30 minutes to reach signal stability after installation.