



www.prosol.co.uk

GOC5230
O2 Sensor

REPLACES: R-17A, R-17AV
R-17AH, AO3, OOA101-1

2018-07

S/N P123456









www.prosol.co.uk

GOC5230 Oxygen Sensor

2018-07

S/N P123456







REPLACES:
R-17A, R-17AV
R-17AH, AO3
OOA101-1

Depending on the ambient conditions the sensor may require up to 30 minutes to reach signal stability after installation. The sensor contains caustic liquid. In case of leakage avoid contact with eyes and skin! Recommended storage temperature: 5°C - 15°C!



www.prosol.co.uk

GOC5223
O2 Sensor

REPLACES: R-22A, R-22AV
R-22AH, AO2, OOA101

2018-07

S/N P123456









www.prosol.co.uk

GOC5223 Oxygen Sensor

2018-07

S/N P123456







REPLACES:
R-22A, R-22AV
R-22AH, AO2
OOA101
Also SUN P/N
7096E406031

Depending on the ambient conditions the sensor may require up to 30 minutes to reach signal stability after installation. The sensor contains caustic liquid. In case of leakage avoid contact with eyes and skin! Recommended storage temperature: 5°C - 15°C!