



Stockholm, September 6, 2021

To whom it may concern,

The Gas Sampling Line H, produced by bluepoint medical GmbH & Co. KG has been used together with the AnaConDa system (Sedana Medical) within a pilot study at 5 European hospitals to evaluate usability of the line.

The parameters evaluated and the preliminary feedback from two of the clinical centres are presented below. The feedback collected so far is positive, both from a functional and from a usability aspect.

Sedana Medical intend to incorporate the gas sampling line to the product offering for all European customers. Furthermore, the line will be used as an investigational device (not FDA approved) in a planned multicentre study in the US.

Average response from 2 clinical centers.

Scale: 4-very pleased, 3-pleased, 2-quite pleased, 1-not pleased	
Overall impression	3
Opening the blister	3.5
Connecting the line to AnaConDa and the gas monitor	4
Reducing number of steps in the set-up	4
Reducing numbers of items when ordering	3
Ability to reduce moisture in the water trap	3
Usability	4
Scale: a-definitely, b-probably, c-probably not, d-definitely not	
Would you recommend this product	b

Yours sincerely,

Peter Fröberg
CTO, Sedana Medical AB, Stockholm